

LAB MONTHLY SUMMARY



Lab Name **Jeba Clinical Laboratory** Lab No **11647**
 Month **January** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 15/01/2021

Date of Report Published : 05/02/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	914	118.27	10.50	12.41	112 mg/dl	-0.51	0.82
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	806	33.34	17.16	5.72	25 mg/dl	-1.46	0.40
3	CREATININE II	Enzymatic Colorimetric II	Any Analyser (Automation / Semi Automation)	80	1.31	18.29	0.24	1.4 mg/dl	0.38	0.05
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	759	2.04	21.65	0.44	2 mg/dl	-0.09	0.03
5	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	696	5.57	16.35	0.91	5.3 mg/dl	-0.30	0.07
6	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	854	115.72	13.16	15.23	128 mg/dl	0.81	1.04
7	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	753	141.84	14.13	20.05	140 mg/dl	-0.09	1.46

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email:clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****

