

**STATUS PASSED**  
**CALIBRATION REPORT**


DEVICE ID:  
METHOD ID: VV200  
ENVIRONMENTAL FACTORS  
TEMP: 20°C  
AIR DENSITY: 0.001200

DESCRIPTION: VAR VOL 20-200µl  
DATE: 21/02/2019  
REL.HUMIDITY: 50%  
CUBIC EXP: 0.0000cm<sup>3</sup>/degc.

SAMPLE				CALIBRATION STATISTICS						
IMPRECISION CV%				INACCURACY E% IMPRECISION CV%						
VOL	NO	WI	VOL	VOL	MEAN	SD	ACTUAL	TARGET	STATUS	
20.0	1	19.90	19.95	20.0	19.95	0.085	0.25	1.0	passed < 0.30	
20.0	2	39.70	19.85	100.0	100.29	0.008	1.00	0.8	passed < 0.30	
20.0	3	59.70	20.06	200.0	199.92	0.476	0.04	0.6	passed < 0.20	
100.0	1	100.00	100.30							
100.0	2	199.99	100.29							
100.0	3	299.99	100.30							
200.0	1	199.90	199.59							
200.0	2	399.90	200.60							
200.0	3	598.00	199.59							

All Pipettes are calibrated at 20°C with degassed double distilled water conforming to DIN12650

Checked By: \_\_\_\_\_

  
 Manager (Quality Control)