

Registered Office:
BioSystems Diagnostics Pvt. Ltd.,
A4, SIPCOT Industrial Park, Iningathukottai,
Thandlam Village, Siperumbudur - 602105. Ph:
Ph: +91 - 44 - 5034 1911 / 12 / 13
E-mail: customersupport@biosystems.in
Website: www.biosystems.in
CIN — U29297TN2003PTCO51492



Date: 09/03/2021

CALIBRATION CERTIFICATE

This is to certify that BTS – 350 Photometer bearing S.No:8017512371 installed at MOLXPRT DIAGNOSTICS PVT. LTD., GREATER NOIDA has been calibrated on 08/03/2021

Calibration Values:

LIGHT INTENSITY TEST

LED	TINT	MAIN	REFERENCE
340	11	947150	200406
405	11	911546	610177
505	25	977210	675854
535	11	955215	470636
560	11	940603	525051
600	20	966488	585027
635	11	957498	540294
670	12	940815	575069

PERISTALTIC PUMP ADJUSTMENT TEST

VALUE	CURRENT	NEW
SAMPLE VOLUME	45877	45877
POSITIONING	2378	2378
PUMP DELAY	2	2

Next calibration is due on 07/03/2022

For BIOSYSTEMS DIAGNOSTICS Pvt. Ltd.,

[Handwritten Signature]



AUTHORISED SIGNATORY

Corporate Office:
BioSystems S.A. Costa Brava 30, 08030 Barcelona (Spain)
Tel +34-93 311 00 00, Fax+34-93 346 77 99
<http://www.biosystems.es>

ISO 13485

LIGHT INTENSITY

08/03/21 14:02:12

LEDS	TINT	MAIN	REFERENCE
340	11	947150	200406
405	11	911546	610177
505	25	977210	675854
535	11	955215	470636
560	11	940603	525051
600	20	966488	585027
635	11	957498	540294
670	12	940815	575069

EXIT

ADJUSTMENT PUMP

08/03/21 13:38:07

TEST VOLUME (μL):

VALUE	CURRENT	NEW
SAMPLE VOLUME	45877	
POSITIONING	2378	
PUMP DELAY	2	

CHECK

MANUAL

AUTOMATIC

EXIT

