



**END CYCLE SUMMARY REPORT
HEMATOLOGY**

Lab Code: 1986

Cycle No. : 09
Jan'2020 - Dec'2020
Date: 28/12/2020

Parameter	Sample Number														
	1		2		3		4		5		6				
	Your value	SDI	Your value	SDI	Your value	SDI	Your value	SDI	Your value	SDI	Your value	SDI			
✓ Hb gm / dl	-	-	-	-	Cancelled				-	-	11.3	-0.8			
✓ WBC × 10 ⁹ / µl	-	-	-	-					-	-	-	-	-	6.1	-2.5
✓ RBC × 10 ⁶ / µl	-	-	-	-					-	-	-	-	-	4.1	0.0
✓ HCT %	-	-	-	-					-	-	-	-	-	35.0	-0.3
✓ MCV fL	-	-	-	-					-	-	-	-	-	84.8	-0.3
✓ MCH pg	-	-	-	-					-	-	-	-	-	78.5	-1.1
✓ MCHC gm / dl	-	-	-	-					-	-	-	-	-	33.6	-0.3
✓ Platelet × 10 ⁹ / µl	-	-	-	-					-	-	-	-	-	298.0	1.0
DLC	Group Mean	PLS-56.8, L7.4-21.7, E38.5-86.7, Mono1.0-3.9, NRBC 0.0-2.0			Stab-0.1-16.6, P69.7-89.5, L5.7-10.9, E1.1-3.0, Mono1.2-3.7, Baso1.2-3.8, NRDC1.8-5.3			Cancelled							
	Your Value	-			-			-							
	Group Mean	Cancelled			L-67.6-96.4, P-1.9-13.6, Mf-8.3-3.1, E-0.8-2.8			B1-0.8-1.7, Pmyelo-0.9-6.2, Myelo-8.0-26.1, Mmyelo-5.5-20.3, S-4.9-19.9, P-19.5-54.4, B-0.8-3.6							
	Your Value	-			-			B1-1, Pmyelo-7, Myelo-7, Mmyelo-3, S-11, P-66, L-3, E-2, NRBC-4							
SDI Value (+/-)	0.0-0.5		0.6-0.9		1.0-2.0		2.1-2.9		>3						
Interpretation	Excellent Performance		Good Performance		Acceptable Performance		Marginal Performance Need Improvement		Unacceptable Performance Needs Urgent Action						

Legends:-

(✓) No Warning	(-) Not Reported
----------------	------------------

Legends:-

(*) Late Result Submission	(S) Reported in other Unit	(*) Not Included in group mean
----------------------------	----------------------------	--------------------------------

Chief Coordinator
[Signature]
Dr. Sanjay Mehrotra

Programme Director
[Signature]
Dr. Bandana Mehrotra

End of Report





RML - Quality Assurance Program (RML - QAP)



Date: 26.10.2020

To,
Reliable Diagnostic Centre
Mr. Akshat Sharma
Plot no.-07, Sdul Colony,
Marwar Hospital, Main Road Bikaner Rajasthan
Ph: 7014867096

Subject: RML Quality Assurance Program 2020-Cycle-Nine (9)

Dear Participant,

We confirm your enrollment in our **Quality Assurance Program Cycle-9, 2020**. Mentioned below are some important details.

1. Your Lab Confidential Code is **1986**
- 2.- Field you opted for

a) Hemat

3. You are requested to mention only your **Lab Code** while submitting the results which can also be mailed to us at rmlresearchfoundation@gmail.com clearly mentioning your lab code in **subject line**.

4. Sample dispatch will have prior information.

5. If you have not received the sample within 5 days of dispatch or there is any discrepancy in the sample received. Kindly send an email urgently. So that corrective action can be taken.

6. Please adhere to mentioned time lines.

"QAP program success is dependent on your cooperation"

Chief PT Coordinator

(Dr. Sanjay Mehrotra)



Address: B-171, Nirala Nagar, Lucknow - 226 020. Ph. : 4034100-130 (30 Lines) 4077180. Fax : (0522)2788555

Email: rmlresearchfoundation@gmail.com

Website: www.rmlpathology.com

Continuous Efforts And Execution Leads To Quality Excellence