

Calibration Certificate

UNIT DESCRIPTION: CLINICAL CENTRIFUGE **MODEL NO.:** R - 303
SR. NO.: ZGGN-13438 **CERTIFICATE NO.:** RMTC - 0548/20
NAME OF CUSTOMER: DR NK SARDANA MD
ADDRESS OF CUSTOMER: SARDANA LABS JALANDHAR
CALIBRATION DATE: 12.07.2020 **CAL. VALIDITY:** 11.07.2021

DETAILS OF EQUIPMENT USED FOR CALIBRATION

NOTE: Tick mark where as applicable.

Digital Lab Thermometer Digital Tachometer Digital Stop Watch
 Auto Ranging R.P.M. Meter Digital Strobe Weight Set
 Digital Temperature Indicator

Sr. No.	Description	Certified By	Instrument Id. No.	Certificate No.	Calibration Date	Validity
01	Digi. Strobe	Autocal	MSM 01	C19120811/P/01	31.12.2019	30.12.2020

OBSERVATIONS:

Sr. No.	Speed Reading on Digital Strobe	Remarks
01	1010	OK
03	2510	OK
05	3830	OK

NOTE: 1. Calibration Traceable to national standards.
 2. Overall deviation is within specified accuracy.
 3. Result is related only to the item calibrated.
 4. Maximum tolerance permissible is $\pm 1^{\circ}\text{C}$ & $\pm 5^{\circ}\text{C}$ whichever is applicable and RPM $\pm 2\%$ FSD and $\pm 5\text{gm}$ to set weight.

REMARKS: ACCEPTED



CALIBRATED BY: VIJAY KADAM

APPROVED BY: MANAGER Q. A.