

CALIBRATION CERTIFICATE

CUSTOMER: LIFE CARE DIAGNOSTIC CENTRE, DHARMAPURI

INSTRUMENT: FULLY AUTO BIOCHEMISTRY ANALYSER

COMPANY	MODEL	SERIAL NO
TULIP DIAGNOSTICS	CORALYZER 200	D31N021

THE REFERENCE OF CALIBRATION: The performance / Adjustments of various subsystems has been calibrated / tested by trained site engineer with pre-designed /calibrated tools provided for the particular subsystems by the company.

Preventive maintenance procedure checklists:

Subsystem Name	
1.Reagent /sample Area A. Sample wheel B.bar code reader C.RFID reader. D.Home Drive belt E.Tray transport Home position.	WORKING
2.Reaction Area A. Photo meter B.cuvette wheel D.Home Drive belt E.Tray transport Home position.	WORKING
3.Transfer Arm/probe A.Arm positioning B.Probe positioning. C.Liquid level sensor.	WORKING
4.Fluid Drawer A. Dilutor pump. B.Waste pump. C.ISE sensor system	WORKING

NOTE:

A. This certificate refers only to the particular item submitted for calibration.

B.The calibration result reported in the certificate is valid at the time of and under the stated condition of the measurement.

5.Heater / Cooler A. Preheat temperature B. Reaction temperature. C. Cooler temperature.	WORKING
6.cover / Door Latches A. Reagent cover B. Cuvette cover C. Drawer	WORKING
Precision Test	✓
System security test	✓
Auto print	✓
A. Auto print all sample result	✓
B. Auto print all patient result	✓
C. Header report.	✓
D.Patient Report test order.	✓
Operator ID	✓
Position for stat	✓
Debug mode	✓
Auto run after inventory	✓
Calibration Result	✓
QC Result	✓
Work list warning	✓

NOTE:

A. This certificate refers only to the particular item submitted for calibration.

B.The calibration result reported in the certificate is valid at the time of and under the stated condition of the measurement.

CALIBRATION CERTIFICATE

CUSTOMER NAME : LIFE CARE DIAGNOSTIC CENTRE

Address: GOUNDER PARK, DHARMAPURI

CONTACT PERSON : Mr.ANBALAGAN

INSTRUMENT : FULLY AUTO BIOCHEMISTRY ANALYSER.

MODEL : CORALYZER 200

S/NO : D31N021

CALIBRATED DATE : 01/01/2021

CALIBRATION DUE DATE: 31/12/2021

THE FOLLOWING SUBSYSTEM HAS BEEN CALIBRATED IN THE INSTRUMENT.

1. REAGENT/SAMPLE AREA
2. REACTION AREA
3. TRANSFER ARM / PROBE
4. FLUID DRAWER
5. COOLER/HEATER
6. COVER / DOOR LATCH
7. PRECISION

TEST

CERTIFICATED BY

G DINESH

Service - Manager



NOTE:

A. This certificate refers only to the particular item submitted for calibration.

B. The calibration result reported in the certificate is valid at the time of and under the stated condition of the measurement.