



Medical Engineering & Services

Things Beyond Thoughts

ISO 9001-2015 certified company

Abhcon Crown, X1/411/21, Ambadi Lane
Kokkalai, Thrissur-21. Ph: 0487 2420367
Email: medengserv@gmail.com
Visit us @www.mescalibration.com



CALIBRATION CERTIFICATE

Certificate No.	MES/20-21/WACHPD/REF/D04	Date of Issue	: 07/01/2021
		Calibration Done On	: 04/01/2021
		Calibration Due On	: 03/01/2022

Customer Name **WOMEN AND CHILDREN HOSPITAL, PALAKKAD.**

DETAILS OF DEVICE UNDER CALIBRATION

Equipment Name	Make / Model	Serial No.	Asset ID	Location
Refrigerator	Whirlpool	-	67090025	ICTC Lab

DETAILS OF MASTER EQUIPMENT USED

SI No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Thermometer	Metravi / DTM 900	10055593	ACS/20-21/I02669-01	23/09/2021
2	Digital Thermo Hygrometer	Equinox / EQ-308CTH	20191602	ACS/MC/20-21/2593-01	18/06/2021

The instruments used for calibration are under valid calibration and are traceable to National Standards.

Environmental Conditions	Temperature	27 °C	Relative Humidity	55 %
--------------------------	-------------	-------	-------------------	------

PERFORMANCE DATA

Display	Measured	Tolerance
1.3 °C	1.4 °C	± 0.3 °C
2.4 °C	2.4 °C	± 0.3 °C
4.2 °C	4.2 °C	± 0.3 °C
7.2 °C	7.1 °C	± 0.3 °C

Calibration Status	Pass	✓	Fail
--------------------	------	---	------

Calibrated By

Approved By

Vivek Mahendran (Service Engineer)



Binoj Thomas (Quality Manager)