



CALIBRATION CERTIFICATE

Certificate No.	MES/20-21/WACHPD/CEN/D03	Date of Issue	: 07/01/2021
		Calibration Done On	: 04/01/2021
		Calibration Due On	: 03/01/2022

Customer Name	WOMEN AND CHILDREN HOSPITAL, PALAKKAD.
---------------	---

DETAILS OF DEVICE UNDER CALIBRATION

Equipment Name	Make / Model	Serial No.	Asset ID	Location
Centrifuge	Rotek	0913	67090027	ICTC Lab

DETAILS OF MASTER EQUIPMENT USED

SI No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Tachometer	Equinox / EQ-801B	S916577	ACS/20-21/102671-01	24/09/2021
2	Digital Thermo Hygrometer	Equinox / EQ-308-CTH	20191602	ACS/MC/20-21/2593-01	18/06/2021

The instruments used for calibration are under valid calibration and are traceable to National Standards.

Environmental Conditions	Temperature	24.5°C	Relative Humidity	55 %
---------------------------------	-------------	--------	-------------------	------


PERFORMANCE DATA


Set Reading	Measured Reading
1	1036 rpm
2	2510 rpm
3	2965 rpm
4	3345 rpm
5	4287 rpm

Calibration Status	Pass	✓	Fail
---------------------------	------	---	------

Calibrated By

Approved By


 Vivick Mahendran (Service Engineer)


 Binoj Thomas (Quality Manager)