



VECTOR BIOTEK PVT. LTD.

Factory & Registered Office : 425, NEW GIDC, KABILPORE, NAVSARI - 396 424, (GUJ.) INDIA
Toll Free No. : 1800-270-1027, FAX : +91-2637-265105, E-mail : vector@beaconindia.com
Website : www.vectorbiotekindia.com, CIN No. : U33112GJ2002PTC040978

Calibration Certificate

Calibration of the **UNICORN 100+** (Fully Auto Biochemistry Analyzer) has been carried out as per the recommendations of the principle company.

Customer Name : **STL DIAGNOSTIC CENTRE, CHENNAI, TN.**

Date of Installation : **05-12-2019**

Calibration Date : **01/03/2021.**

Next Calibration Due On: **01/09/2021.**

Tested By:



Checked By:





VECTOR BIOTEK PVT. LTD.

Factory & Registered Office : 425, NEW GIDC, KABILPORE, NAVSARI - 396 424. (GUJ.) INDIA
Toll Free No. : 1800-270-1027, FAX : +91-2637-265105, E-mail : vector@beaconindia.com
Website : www.vectorbiotekindia.com, CIN No. : U33112GJ2002PTC040978

Performance Qualification

MODEL: UNICORN 100+

Sr. No CA15191100117

Version: 1.0.14

Date: 01-03-2021

PQ is performed at

STL DIAGNOSTIC CENTRE, CHENNAI, TN.

According to the assay results that are comprised in this document the UNICORN 100+ system can be approved for the routine operation.

System released for routine operation.

PERFORMANCE CALIBRATION			
LOT No: MB092004	Exp Date: 2022-02		
TEST	TARGET	RANGE	RESULT
ALB	4.33	3.9-4.76	4.29
ALKP	190	157.5-222.5	196.20
CHOL	190	157.5-222.5	192.83
CRE	3.6	2.99-4.21	3.11
GLU	257	213.5-300.5	299.05
SGPT	113	93.5-132.5	119
TG	209	173-245	204
TP	7.42	6.67-8.17	7.22
UREA	140	118-162	146
URIC ACID	9.51	7.89-11.13	9.07

Tested By:



Checked By:

