


## LAB MONTHLY SUMMARY



Lab Name **SURYA DIAGNOSTIC CENTRE** Lab No **14213**  
 Month **January** Year **2021**  
 Constituent Group **Chemistry II**



Date of Result Entered : 11/01/2021

Date of Report Published : 05/02/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE II	GOD-POD II	Biosystems	255	119.68	9.01	10.79	124 mg/dl	0.40	1.35
2	UREA II	Others ( Urease Berthelot / Ned Dye ) II	Any Analyser (Automation / Semi Automation )	806	33.34	17.16	5.72	28 mg/dl	-0.93	0.40
3	CREATININE II	Jaffes End point II	Any Analyser (Automation / Semi Automation )	532	1.32	12.99	0.17	1.4 mg/dl	0.47	0.01
4	T.BILIRUBIN II	Diazonium Salt ( Colorimetric ) / Jendrassik II	Any Analyser (Automation / Semi Automation )	759	2.04	21.65	0.44	1.3 mg/dl	-1.68	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Biosystems	209	5.46	10.49	0.57	5.7 g/dl	0.42	0.08
6	ALBUMIN II	BCG - Colorimetric II	Biosystems	204	3.33	9.62	0.32	3.5 g/dl	0.53	0.04
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Biosystems	202	5.61	14.10	0.79	6.1 mg/dl	0.62	0.11
8	CHOLESTROL II	CHOD-PAP II	Biosystems	240	115.43	10.70	12.35	125 mg/dl	0.78	1.59
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Biosystems	208	143.79	12.33	17.73	142 mg/dl	-0.10	2.46

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between $\pm 1.0$ to $\pm 2.0$	Good.
Between $\pm 2.0$ to $\pm 3.0$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email:clinq@cmcvellore.ac.in

Contact Number: 0416-2283102

*Parvati Christudoss*

**Dr. Pamela Christudoss**

**CMC EQAS Co-Ordinator**

**Christian Medical College, Vellore**

**\*\*\*\*\* End of Report \*\*\*\*\***