



**Labtech
Healthcare**

Labtech Health Care India Pvt. Ltd.

193/SP-6 & 7, Kumbha Marg, Near IRI Circle,
Pratap Nagar, Sangarner, Jaipur-302033 INDIA
+91-141-2712685, +91 98296 25650

CALIBRATION CERTIFICATE

Customer Name	RELIABLE DIAGNOSTIC CENTRE SHRI HOSPITAL		
Department	Department of Haematology-Laboratory Medicine		
Address	4, Vishnupuri, Jagatpura Road, Jaipur, pincode-302017		
	Date of Calibration	18.03.2021	
Instrument	<i>Mindray BC-5130 ANALYZER</i>	Next Calibration Date	17.03.2022
Serial No	TR-98003386	Calibrator Lot Number	PLUS-0221
		Calibrator Expiry date	05-07-2018

	<u>Parameter</u>	<u>Target</u>	<u>Range</u>	<u>Mean of Measured Values</u>	<u>New Factor</u>	<u>Old Factor</u>
<u>Calibration Values</u>	WBC x 10 ⁹ /L	8.49	8.29-8.69	8.28	108.43	107.42
	RBC x 10 ¹² /L	4.45	4.35-4.455	4.38	103.86	103.67
	HGB g/L	133	131-135	13.1	103.43	103.10
	MCV fl	97.2	95.2-99.2	99.3	91.61	92.09
	PLT x 10 ⁹ /L	262	247-277	252	102.55	101.85

This is to certify that the Mindray BC 6800 PLUS Hematology Analyser Sr.No: TR-98003386 system, installed at RELIABLE DIAGNOSTIC CENTRE SHRI HOSPITAL, Calicut has been calibrated using the standard reference material SC CAL PLUS-0618 Calibrator. Traceability of SC CAL PLUS Calibrator material from the manufacturer is attached along with this certificate.

Date: 19-3-2021

Approved by

For Labtech Healthcare India Pvt. Ltd.
Jaipur (Raj.)



Branch Office: 44, Sudama Nagar, Opp. Glass Factory Tonk Road, Jaipur-302015 INDIA

info@labtechhealthcare.com www.labtechhealthcare.com