

**CALIBRATION CERTIFICATE**

Customer Name	TOMAR FOUNDATION PATHOLOGY,VIHAR.
Instrument Name	TURBOCHEM 100
Installation Date	23/04/2016
Serial Number	4621-1166
Calibrated On	16/07/2020
Next Calibration Due On	15/07/2021

**Filter Voltages**

The Photometric Voltages should be within 2V to 10V

<b>FILTERS(nm)</b>	<b>VOLTS(v)</b>
340	7.1
405	7.4
505	7.3
545	7.3
580	7.2
630	7.3

**Pipetting Check**

Requirement : The C.V. Should be less than 2% for Medium & Large, Less than 3% for Small.

**Self Test -Passed**

Type	Target	Mean	SD	CV
<b>SMALL</b>	0.3560	0.3882	0.0088	2.2685
<b>MEDIUM</b>	1.7300	1.6963	0.0037	0.2198
<b>LARGE</b>	2.6720	2.5643	0.0244	0.9500

**Incubation Temperature**

Range : 37° C + 0.5

On Board display the Temperature is 37.0 degree stable

On complete testing of the unit we found that the unit is working satisfactorily.And the same can be used to carry out the necessary functions.

For CPC Diagnostics (P) Ltd.,



Authorised Signatory



# INVOICE

**CPC Diagnostics Pvt. Ltd., Chennai**  
 New No.324, Old No.237/1, Khadi Gramodyog Complex,  
 Anna Shanmugam Salai, Gopalapuram,  
 Chennai  
 600086

**Registered Office**  
 9 & 10, Gokul Towers, No-9, V Floor, C.P. Ramaswamy Road,  
 Alwarpet, Chennai-600018.

**Branch Office -1 :**  
 Khadi Gramodyog, New No.324, (Old No.237/1) Avvai  
 Shanmugam Salai, Gopalapuram, Chennai-600086,  
 Phone:23460161-64, Fax :23460165-66.

**Branch Office -2 :**  
 2555, Babu Jagajeevanram Street, Gerugambakkam,  
 Kundrathur Road, Chennai - 602101

Tamil Nadu  
 Dist: 431B/MIII/20B, 4204 MIII/21B Dt. 15-10-07  
 GSTIN: 32AAACC3142N1ZV

<b>Bill To,</b> COMAR FOUNDATION PATHOLOGY ADD:17 A POCKET 10, OPP BSES OFFICE MAYUR VIHAR, NEW DELHI 110091 State : Delhi GSTIN : UR DELHI	<b>Ship To,</b> COMAR FOUNDATION PATHOLOGY ADD:17 A POCKET 10, OPP BSES OFFICE MAYUR VIHAR, NEW DELHI 110091 State : Delhi GSTIN : UR DELHI	<b>Invoice No. &amp; Date</b> AMC/CMC132/20-21- 21/07/2020	<b>Docket / LR No. &amp; Date</b>
		<b>Delivery Document No.</b>	<b>No. of Parcels</b>
		<b>Your Order No. &amp; Date</b>	
		<b>Transportation Mode</b>	

CODE	Description of goods/ Services	HSN	Batch No	Exp.Date	Qty	Unit	Unit Price	Amount	Dis count %	Taxable Value	CGST		SGST		IGST		Total Amount
											Rate %	Amt	Rate %	Amt	Rate %	Amt	
AMC100	AMC TURBO CHEM 100 AMC CHARGES FOR TC100 S.NO. 4621 1166 PERIOD FROM 16-07-2020 TO 15-07- 2021	9987			1	Nos	27000.00	27000.00	0%	27000.00					18	4860.00	31860.00
<b>Total :</b>					1.00			27000.00		27000.00					4860.00	31860.00	

**Bill Amount : INR THIRTY ONE THOUSAND EIGHT HUNDRED SIXTY ONLY**

HSN	TAXABLEAMT	QTY	CGST	SGST	IGST	TOTALAMT
9987	27000.00	1.00			18	4860.00
<b>Total:</b>	<b>27000.00</b>	<b>1</b>				<b>4860.00</b>

**Bank Details**

A/C Name : CPC DIAGNOSTICS PRIVATE LIMITED  
 A/C No : Cash Credit Account No. 0344609004  
 Bank Name & Branch : CITI Bank N.A. Chennai  
 IFSC Code : CITI0000003  
 MICR Code No : 600037002  
 PAN NO : AAACC3142N

**Terms and Conditions :**

1. Goods once sold cannot be taken back
2. Interest @ 18% is chargeable for payment after due date
3. The sale is subject to the terms and conditions mentioned in the invoice.
4. All disputes which may arise in connection with this terms and conditions of sale shall be resolved by recourse to Arbitration in Chennai.
5. Any dispute shall be referred to a sole arbitrator to be appointed by CPC Diagnostics Pvt. Ltd
6. Only Courts in Chennai shall have the exclusive jurisdiction over all issues relating to this terms and conditions sale.

For CPC Diagnostics Pvt. Ltd.,



**Authorised Signatory**