



# EURO CALIBRATION TECHNIQUE

No. 387, IInd Floor, TIC Road, Varadharajapuram,  
Ambattur, Chennai-600 053. Tamil Nadu

Tel: 044-48521288 Mobile: 9486195184, 9176032184

E-mail: eurocalibration@gmail.com web: www.eurocalibration.com



## CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2021/0237- 01

ULR Number : CC2519 21 0 00002417F

Customer Name and Address

M/s. PPTCT JIPMER,  
Women & Children Block,  
Department Of Gynaecology,  
PPTCT JIPMER, Puducherry- 605 006.

Date of Issue : 13-Mar-2021

Pages : 1 of 1

SRF Number : 2021/0237

Receipt Date : 11-Mar-2021

Date of Calibration : 11-Mar-2021

Suggested calibration Due Date : 10-Mar-2022

Calibrated at : Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-M-09

Equipment Under Calibration

Physical Condition : Good

Name : Centrifuge

Make : Remi

Model : R-8C

S.No : ZCGN-19263

ID No. : --

Range : 0 to 5250 RPM

Resolution : 1 RPM

Accuracy : As Per Manual

Location : OGOPD (Room-6)

Environmental Conditions

Temperature : 26.8 °C

Relative Humidity : 56 %RH

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Digital Tachometer	S622194	ECT 2020/INT-24	08.11.2021

Calibration Results

Parameter - Mechanical /Acceleration and Speed

S.No	Test Reading RPM	Standard Reading in RPM			Standard Reading (Mean) RPM	Error RPM	Error Allowed ±RPM	Uncertainty ±RPM
		1	2	3				
1	98	97.8	97.6	97.7	97.70	0.30	--	3.90
2	495	494.6	494.6	494.7	494.63	0.37	--	3.90
3	2007	2010	2011	2013	2011.33	-4.33	--	4.14
4	4450	4457	4459	4460	4458.67	-8.67	--	4.14
5	5011	5022	5026	5027	5025.00	-14.00	--	4.56

Checked By



Authorized Signatory  
K.ABDUL VAJITH / R.SATHYAKALA

REMARKS: 1. Calibration certificate traceable to National / international standards which realize the units of measurement according to the international system of units (SI).  
2. This certificate shall not be reproduced except in full without return approval of Euro Calibration Technique. 3. The Calibration results reported in the certificate as valid at the time of and under the stated condition of measurement for the particular equipment. 4. Next Calibration due date mentioned as per the customer requirement. 5. The measurement confidence level is 95% for a coverage factor k=2. 6. Calibration certificate issued for weights & measure parameter i.e mass, balance, volumetric equipments are for scientific or industrial purpose only.