

## APEX QUALITY CERTIFICATION SERVICES PVT. LTD.

Add.: 161, Kasturba Nagar, Nirman Nagar, Jaipur-302 019 Ph.: 0141-4036776, 9414079483, 9636293333

Email: apex.gcs@gmail.com



NABL ACCREDITED LAB Cert. No.: CC-2298

CALIBRATION CERTIFICATE

ULR: CC229820000002645F		Page 01 of 01
CUSTOMER'S DETAILS	CERTIFICATE NO.:	AQCS/20121901.0001
M/S SHRI HOSPITAL	Date of Issue of Certificate	19-12-2020
4-5, Vishnupuri, Jagatpura Road,	Date of Calibration	19-12-2020
Jaipur (Rajasthan)	Recommended due date	18-12-2021

Service Request	No.	201217/01		INSTRUMENT DESC	RIPTION	
Service Request	Date	17-12-2020		Nomenclature	Centrifuge Ma	chine
DUC Receipt Dat	е	19-12-2020		Make	Remi	
DUC Condition		OK		Model	MEDICOPLUS	
Cal. Performed A	t	Site		Sr. No. / Ident. No.	ZECU-09521/S	SH/LAB/CFG/01
Procedure Follow	ed	CP/RM/02		Range	0-4000	RPM
Environmental C	ondition			Least Count	10	RPM
Temperature	23±5°C	Rel. Hum.	50±10%	Location	Lab	

## STANDARD EQUIPMENT USED:

Nomenclature	Range	Uncertainty (±)	Identification No.	Traceability Cert No.	Valid up to
Digital Tachometer	10 to 99,999	5 to 26	S999205	C&IJ/CAL/20-02/160	25-02-2021

## CALIBRATION RESULT

Measured Value of RPM on Std.	Measured Value of RPM on DUC	Error in RPM	Expanded Uncertainty (± % of rdg.) in RPM
600.6	600	-0.6	2.37
1502	1500	-2	2.37
3053	3050	-3	1.04
4053	4050	-3	1.04
5224	5220	-4	1.04

## REMARKS:

- 1. Expanded uncertainty of Measurement at approximately 95 % confidence level for coverage factor k = 2 is as given above.
- 2. The result reported in this certificate is valid at the time of and under the stated conditions of Measurement for the particular item and instrument identified above.
- 3. This certificate shall not be reproduced, except in full, unless a written permission for the publication of an approved abstract has been obtained from the Director, Apex Quality Certification Services Pvt. Ltd. Jaipur

4. Format Number.: F-06/00

Calibrated by
(Sachin Yadav)

Approved by

(R. R. Yadav) Quality Manager