



APEX QUALITY CERTIFICATION SERVICES PVT. LTD.



Add.: 161, Kasturba Nagar, Nirman Nagar, Jaipur-302 019
Ph.: 0141-4036776, 9414079483, 9636293333
Email: apex.qcs@gmail.com

CALIBRATION CERTIFICATE

CERTIFICATE NO.: AQCS/21031301.10003		Date of Calibration : 13/03/2021		Page 01 of 01
		Recommended due date : 12/03/2022		
CUSTOMER'S DETAILS		INSTRUMENT DESCRIPTION		
M/S RELIABLE DIAGNOSTIC CENTER C/O SHRI HOSPITAL 4, Jagatpura Road, Jaipur-302017		Nomenclature : Temp. Indicator with sensor (Refrigerator) Make : R-tek Model : RT-608 Sr. No. : RT-28079 Ident. No. : NA Range : 2-8°C LC : 0.1°C Location : -- Cal. Performed At : Site		
Service Request No. : 210310/01 Dated: 10/03/2021				
STANDARD EQUIPMENT USED:				
Nomenclature	Identification No.	Traceability Certificate No.	Valid up to	
PT100 RTD with indicator & Source	11339/16A046131	TL/020/319.2.1	28/05/2021	
ENVIRONMENTAL CONDITIONS:			PROCEDURE FOLLOWED:	
Temperature : 25 ± 4°C Relative Humidity : 50± 10%			CP/TH -08	
CALIBRATION RESULT				
Measured Temp. on STD °C	Measured Temp. on DUC °C	Error °C	Expanded uncertainty (±°C)	
2.15	2.3	0.15	2.5	
3.01	3.2	0.19	2.5	
5.15	5.4	0.25	2.5	
6.27	6.6	0.33	2.5	
8.59	8.0	0.41	2.5	
REMARKS :				
1. Expanded uncertainty of Measurement at approximately 95 % confidence level for coverage factor k = 2 is as given above.				
2. The result reported in this certificate is valid at the time of and under the started conditions of Measurement for the particular item and instrument identified above.				
3. This certificate shall not be reproduced, except in full, unless a written permission for the publication of an approved abstract has been obtained from the Director, Apex Quality Certification Services Pvt. Ltd. Jaipur				
4. Format Number.: F-06/00				
Calibrated by (Sachin Yadav)		Approved by (R. R. Yadav) Quality Manager		

