



# APEX QUALITY CERTIFICATION SERVICES PVT. LTD.



Add.: 161, Kasturba Nagar, Nirman Nagar, Jaipur-302 019  
Ph.: 0141-4036776, 9414079483, 9636293333  
Email: apex.qcs@gmail.com

## CALIBRATION CERTIFICATE

Page 01 of 01

CERTIFICATE NO.: AQCS/20121901.10001		Date of Calibration : 19/12/2020		
		Recommended due date : 18/12/2021		
<b>CUSTOMER'S DETAILS</b>		<b>INSTRUMENT DESCRIPTION</b>		
M/S SHRI HOSPITAL 4-5, Vshnupuri, Jagatpura Road, Jaipur (Rajasthan)		Nomenclature : <b>Microscope</b> (Magnification) Make : Labomed Model : -- Sr. No. / Ident. No. : -- / SH/LAB/MS/01 Magnification : 4X, 10X, 40X, 100X Location : LAB Cal. Performed At : Site		
Service Request No. : 201217/01 Dated: 17/12/2020				
<b>STANDARD EQUIPMENT USED:</b>				
<b>Nomenclature</b>	<b>Identification No.</b>	<b>Traceability Certificate No.</b>	<b>Valid up to</b>	
Set of Gauge blocks	3079	CC247218000000887F	29/12/2020	
Digital Vernier Caliper	AQCS-VC-01	AQCS/20021901.002	18/02/2021	
<b>ENVIRONMENTAL CONDITIONS:</b>		<b>PROCEDURE FOLLOWED:</b>		
Temperature : 20 ± 2°C Relative Humidity : 50 ± 10%		CP/MS -01		
<b>CALIBRATION RESULT</b>				
<b>Size</b>	<b>Magnification</b>		<b>Standard reading mm</b>	<b>Magnification Measured</b>
1mm	4X	Projected piece	3.99	3.99X Magnification
1mm	10X	do	10.00	10.00 X Magnification
1mm	40X	do	40.00	40.00 X Magnification
1mm	100X	do	99.99	99.99 X Magnification
1mm	4X	Eye Piece	4.00	4.00 X Magnification
1mm	10X	do	10.00	10.00 X Magnification
1mm	40X	do	40.00	40.00 X Magnification
1mm	100X	do	99.99	99.99 X Magnification
<b>REMARKS :</b>				
1. Expanded uncertainty of Measurement at approximately 95 % confidence level for coverage factor k = 2 is ± 14.1µm.				
2. The result reported in this certificate is valid at the time of and under the started conditions of Measurement for the particular item and instrument identified above.				
3. This certificate shall not be reproduced, except in full, unless a written permission for the publication of an approved abstract has been obtained from the Director, Apex Quality Certification Services Pvt. Ltd. Jaipur.				
4. Format Number.: F-06/00				
Validated by  (Sachin Yadav)		Approved by  ( R. R. Yadav ) Quality Manager		