



EURO CALIBRATION TECHNIQUE

No. 387, IInd Floor, TIC Road, Varadharajapuram,
Ambattur, Chennai-600 053. Tamil Nadu
Tel: 044-48521288 Mobile: 9486195184, 9176032184
E-mail: eurocalibration@gmail.com web: www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2021/0237- 08

ULR Number : CC2519 21 0 00002424F

Customer Name and Address

M/s. Indira Gandhi Government General Hospital
And Post Graduate Institute,
II Floor, ICTC IGGGH, Department of Microbiology,
Victor Simonel Street, Puducherry - 605 001.

Date of Issue : 13-Mar-2021

Pages : 1 of 1

SRF Number : 2021/0237

Receipt Date : 11-Mar-2021

Date of Calibration : 11-Mar-2021

Suggested calibration Due Date : 10-Mar-2022

Calibrated at : Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-08

Equipment Under Calibration

Physical Condition : Good

Name : Refrigerator

Make : Elanpro

Model : ECG374

S.No : A091200009

ID No. : GH/MIC/ICTC/RE/01

Range : 2 to 8 °C

Resolution : NA

Accuracy : As Per Manual

Location : Serology 227-A

Environmental Conditions

Temperature 25.7 °C

Relative Humidity 55 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	RTD Sensor	1720	ECT 2020/INT-36	09.11.2021
2	RTD Calibrator	091109171	20/52-02	09.06.2021

Calibration Results

Parameter - Thermal / Temperature

S.No	Test Reading °C	Standard Reading in °C			Standard Reading (Mean) °C	Uncertainty ±°C
		1	2	3		
1	2 to 8	6.7	6.7	6.7	6.70	0.26

Checked By



Authorized Signatory
K.ABDUL VAJITH / R.SATHYAKALA

REMARKS: 1. Calibration certificate traceable to National / international standards which realize the units of measurement according to the international system of units (SI).
2. This certificate shall not be reproduced except in full without return approval of Euro Calibration Technique. 3. The Calibration results reported in the certificate as valid at the time of and under the stated condition of measurement for the particular equipment. 4. Next Calibration due date mentioned as per the customer requirement. 5. The measurement confidence level is 95% for a coverage factor k=2. 6. Calibration certificate issued for weights & measure parameter i.e mass, balance, volumetric equipments are for scientific or industrial purpose only.