



EURO CALIBRATION TECHNIQUE

No. 387, IInd Floor, TIC Road, Varadharajapuram,
Ambattur, Chennai-600 053. Tamil Nadu
Tel: 044-48521288 Mobile: 9486195184, 9176032184
E-mail: eurocalibration@gmail.com web: www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2021/0237- 04
ULR Number : CC2519 21 0 00002420F

Date of Issue : 13-Mar-2021
Pages : 1 of 1
SRF Number : 2021/0237
Reciept Date : 11-Mar-2021
Date of Calibration : 11-Mar-2021
Suggested calibration Due Date : 10-Mar-2022
Calibrated at : Onsite

Customer Name and Address

M/s. Rajiv Gandhi Women & Children Hospital,
PPTCT RGGW, Department of Gynaecology
Ellaipaillaichavady,
Puducherry- 605 005.

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-08

Equipment Under Calibration

Physical Condition : Good

Name : Refrigerator
Make : Electrolux (Kelvinator)
Model : --
S.No : --
ID No. : PPTCT/RGGWCH/PDY/INS NO.1
Range : 2 to 8 °C
Resolution : NA
Accuracy : As Per Manual
Location : Gynaecology OPD

Environmental Conditions

Temperature 25.6 °C
Relative Humidity 56 %RH
Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	RTD Sensor	1720	ECT 2020/INT-36	09.11.2021
2	RTD Calibrator	091109171	20/52-02	09.06.2021

Calibration Results

Parameter - Thermal / Temperature

S.No	Test Reading °C	Standard Reading in °C			Standard Reading (Mean) °C	Uncertainty ± °C
		1	2	3		
1	2 to 8	5.8	5.8	5.9	5.83	0.26



(Signature)
Authorized Signatory
K.ABDUL VAJITH / R.SATHYAKALA

(Signature)
Checked By

... National / International standards which realize the units of measurement according to the international system of units (SI) ... certificate