



EURO CALIBRATION TECHNIQUE

No. 387, IInd Floor, TIC Road, Varadharajapuram,
Ambattur, Chennai-600 053. Tamil Nadu
Tel: 044-48521288 Mobile: 9486195184, 9176032184
E-mail: eurocalibration@gmail.com web: www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2020/0707- 05
ULR Number : CC2519 20 0 00006566F

Date of Issue : 3-Oct-2020

Pages : 1 of 1

Customer Name and Address

SRF Number : 2020/0707

M/s State Reference Laboratory For HIV,
Department of Microbiology,
Jawaharlal Institute Of Post Graduate,
Medical Education And Research, Jipmer Campus Road,
Gorimedu Dhanvantari Nagar, Puducherry-605 006.

Receipt Date : 1-Oct-2020

Date of Calibration : 1-Oct-2020

Suggested calibration Due Date : 30-Sep-2021

Calibrated at : Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-T-08

Equipment Under Calibration

Physical Condition : Good

Name : Refrigerator

Make : Elanpro

Range : 2 to 8°C

Model : ECG374

Resolution : 1 °C

S.No : A09120023

Accuracy : ± 1 °C

ID No. : JIP/ICTC/HIV/RF/03

Location : State Reference Laboratory

Environmental Conditions

Temperature 24.5 °C

Relative Humidity 63 %RH

Temperature Scale : ITS 90

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No.	Certificate number	Validity
1	RTD Sensor	1711	2019-20/11/INT-32	17.11.2020
2	RTD Calibrator	11110002959	19/127-02	30.10.2020

Calibration Results

Parameter - Thermal / Temperature

S.No	Test Reading °C	Standard Reading in °C			Standard Reading (Mean) °C	Error °C	Error Allowed ±°C	Uncertainty ±°C
		1	2	3				
1	2	2.3	2.3	2.3	2.30	-0.30	1.00	0.69
2	4	4.2	4.2	4.4	4.27	-0.27	1.00	0.70
3	6	5.6	5.6	5.6	5.60	0.40	1.00	0.69
4	8	7.4	7.4	7.4	7.40	0.60	1.00	0.69

[Signature]
Checked By



[Signature]
Authorized Signatory
K.ABDUL VAJITH / R.SATHYAKALA

REMARKS: 1. Calibration certificate traceable to National / international standards which realize the units of measurement according to the international system of units (SI).
2. This certificate shall not be reproduced except in full without return approval of Euro Calibration Technique. 3. The Calibration results reported in the certificate
as valid at the time of and under the stated condition of measurement for the particular equipment. 4. Next Calibration due date mentioned as per the customer
requirement. 5. The measurement confidence level is 95% for a coverage factor k=2. 6. Calibration certificate issued for weights & measure parameter i.e mass,
balance, volumetric equipments are for scientific or industrial purpose only.