



EURO CALIBRATION TECHNIQUE

No. 387, IInd Floor, TIC Road, Varadharajapuram,
Ambattur, Chennai-600 053. Tamil Nadu
Tel: 044-48521288 Mobile: 9486195184, 9176032184
E-mail: eurocalibration@gmail.com web: www.eurocalibration.com



CERTIFICATE OF CALIBRATION

Certificate Number : ECT 2020/0707- 01
ULR Number : CC2519 20 0 00006562F

Date of Issue : 3-Oct-2020

Pages : 1 of 1

Customer Name and Address

M/s State Reference Laboratory For HIV,
Department of Microbiology,
Jawaharlal Institute Of Post Graduate,
Medical Education And Research, Jipmer Campus Road,
Gorimedu Dhanvantari Nagar, Puducherry-605 006.

SRF Number : 2020/0707

Reciept Date : 1-Oct-2020

Date of Calibration : 1-Oct-2020

Suggested calibration Due Date : 30-Sep-2021

Calibrated at : Onsite

Customer Reference : Work Order

Calibration Procedure : ECT/WI-M-09

Equipment Under Calibration

Name : Centrifuge
Make : Remi
Model : R-23
S.No : VCPN-5132
ID No. : JIP/SRL/HIV/CEN/01

Physical Condition : Good

Range : Max 6000 RPM

Resolution : 10 RPM

Accuracy : As Per Manual

Location : State Reference Laboratory

Environmental Conditions

Temperature 25.2 °C Relative Humidity 52 %RH

Standard Equipment used for Calibration

S.No	Standard Instrument	Serial No	Certificate number	Validity
1	Digital Tachometer	S622194	2019-20/11/INT-01	12.11.2020

Calibration Results

Parameter - Mechanical / Acceleration and Speed

S.No	Test Reading RPM	Standard Reading in RPM			Standard Reading (Mean) RPM	Error RPM	Error Allowed ±RPM	Uncertainty ±RPM
		1	2	3				
1	1000	1001	1002	1002	1001.67	-1.67	--	7.53
2	2000	1998	1998	1998	1998.00	2.00	--	7.51
3	3000	3002	3002	3002	3002.00	-2.00	--	7.51
4	4000	4002	4002	4002	4002.00	-2.00	--	7.51

V. Vijay
Checked By



K. Abdul Vajith
Authorized Signatory
K.ABDUL VAJITH / R.SATHYAKALA

REMARKS: 1. Calibration certificate traceable to National / international standards which realize the units of measurement according to the international system of units (SI).
2. This certificate shall not be reproduced except in full without return approval of Euro Calibration Technique. 3. The Calibration results reported in the certificate as valid at the time of and under the stated condition of measurement for the particular equipment. 4. Next Calibration due date mentioned as per the customer requirement. 5. The measurement confidence level is 95% for a coverage factor k=2. 6. Calibration certificate issued for weights & measure parameter i.e mass, balance, volumetric equipments are for scientific or industrial purpose only.