

VIEW LAB MONTHLY SUMMARY

Lab Name	ALPHA LIFECARE	Details About Robust Analysis								
Lab No	8870	Detail About Monthly Summary								
Month	February	Detail about SDI								
Year	2021									
Constituent Group	Chemistry II									
Click on the analyte to view Graphical Data										
All Analyser Result Print										
Date of Result Entered : 24/02/2021										
Date of Report Published : 06/03/2021										
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	747	251.19	10.64	26.72	317 mg/dl	2.46	1.95
2	UREA II	Urease UV / GLDH II	Transasia / Erba	584	54.62	11.45	6.25	112 mg/dl	9.18	0.52
3	CREATININE II	Jaffes End point II	Any Analyser (Automation / Semi Automation)	566	2.74	13.49	0.37	6 mg/dl	8.81	0.03
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	532	1.22	19.52	0.24	0.9 mg/dl	-1.34	0.02
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	580	5.56	10.15	0.56	7 g/dl	2.55	0.05
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	570	3.31	9.68	0.32	3.8 g/dl	1.53	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	576	3.83	16.03	0.61	4.1 mg/dl	0.44	0.05
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	655	110.72	11.41	12.63	121 mg/dl	0.81	0.99
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	504	90.04	16.84	15.16	92 mg/dl	0.13	1.35
SDI Range		Interpretation								
Within -1.0 to +1.0		Excellent.								
Between ±1.0 to ±2.0		Good.								
Between ±2.0 to ±3.0		Accept with caution. Warning Signal.								
Beyond ±3.0		Unacceptable performance. Action Signal.								

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

