

LAB MONTHLY SUMMARY



Lab Name **SOBHI DIAGNOSTIC CENTRE** Lab No **2739**
 Month **October** Year **2020**
 Constituent Group **Chemistry II**



Date of Result Entered : 12/10/2020

Date of Report Published : 09/11/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Robonik (Semi automatic / Fully Automatic)	411	165.51	11.40	18.87	199.6 mg/dl	1.81	1.86
2	UREA I	Urease Berthelot	Robonik (Semi automatic / Fully Automatic)	188	104.25	16.06	16.74	84.5 mg/dl	-1.18	2.44
3	CREATININE I	Jaffes End point	Any Analyser	358	4.52	13.11	0.59	3.2 mg/dl	-2.23	0.06
4	T.BILIRUBIN I	Diazonium salt (Colorimetric) / Jendrassik	Robonik (Semi automatic / Fully Automatic)	177	3.69	19.79	0.73	2.2 mg/dl	-2.04	0.11
5	T-PROTEIN I	Biuret - Colorimetric	Robonik (Semi automatic / Fully Automatic)	149	5.16	11.80	0.61	5.8 g/dl	1.05	0.10
6	ALBUMIN I	BCG - colorimetric	Robonik (Semi automatic / Fully Automatic)	146	3.01	10.73	0.32	4.3 g/dl	4.03	0.05
7	CHOLESTEROL I	CHOD-PAP	Robonik (Semi automatic / Fully Automatic)	366	102.64	11.50	11.81	134.7 mg/dl	2.71	1.23
8	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Robonik (Semi automatic / Fully Automatic)	194	176.93	12.27	21.70	177.7 mg/dl	0.04	3.12

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****



SOBHI *Diagnostic Centre*

A/c Computerised
Tamilnadu Govt Regd. No : 02531/84

✓ 18/34, Perumal Coil Street,
VALLIOOR - 627 117, Tirunelveli Dist.
☎ : 9487959009, ☎ : 04637 - 220798

Branch : Main Road, North Bazar,
ANJUGRAMAM - 629 401, Kanyakumari Dist.
☎ : 8903966499 ☎ : 04652 - 267125




Govt. of Tamil Nadu, Directorate of Medical and Rural Health Services CEA Permission No.: TNVLALL 2020000 8372 / 2019.
Tamil Nadu Pollution Control Board Permission No.: 19 BA, D 28501750/2019, 20GAD 28504164/2019

An ISO 2001:2015 Certified, ISAC Accreditation and
CMC Vellore Monthly Quality Control Lab. Or.No. : CMC/QC/2739/08, ISO No. : ISO915181388

RESOLVING QC FAILURE

Steps taken to resolve out of range values for the month of October

1. The control materials were retested
2. Fresh reagent were prepared and samples were re-run
3. Reagent slots were reviewed
4. Sample storage and handling was reviewed
5. pipettes and instruments were checked



John cast

SOBHI DIAGNOSTIC CENTRE
18/34, Perumalcoil Street
Vallioor - 627 117
Tirunelveli District



We pray for your speedy recovery
MJF Ln. J. Sobitha Raj, M.A., B.Ed., M.A., D.M.L.T.,
Mrs. P. Sarojini Sobitha Raj, M.A., D.M.L.T., M.I.P.H.&H

