

LAB MONTHLY SUMMARY



Lab Name **SOBHI DIAGNASTIC CENTER** Lab No **11625**
 Month **November** Year **2020**
 Constituent Group **Chemistry II**



Date of Result Entered : 19/11/2020

Date of Report Published : 05/12/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Robonik (Semi automatic / Fully Automatic)	391	372.08	10.95	40.76	341 mg/dl	-0.76	4.12
2	UREA I	Urease UV / GLDH	Robonik (Semi automatic / Fully Automatic)	106	57.83	10.07	5.82	60 mg/dl	0.37	1.13
3	CREATININE I	Jaffes End point	Any Analyser	352	7.76	12.12	0.94	6.7 mg/dl	-1.13	0.10
4	T.BILIRUBIN I	OTHERS (Any Other Principles / Methods)	Any Analyser	163	1.77	19.89	0.35	2.4 mg/dl	1.79	0.06
5	T-PROTEIN I	Biuret - Colorimetric	Robonik (Semi automatic / Fully Automatic)	158	5.35	11.39	0.61	5.9 g/dl	0.90	0.10
6	ALBUMIN I	BCG - colorimetric	Robonik (Semi automatic / Fully Automatic)	151	3.21	9.64	0.31	2.5 g/dl	-2.30	0.05
7	CHOLESTEROL I	CHOD-PAP	Robonik (Semi automatic / Fully Automatic)	362	111.52	11.57	12.91	105 mg/dl	-0.51	1.36
8	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Robonik (Semi automatic / Fully Automatic)	201	156.43	14.98	23.43	119 mg/dl	-1.60	3.31

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****



GOD IS GREAT

SOBHI Diagnostic Centre

A/c Computerised

Tamilnadu Govt Regd. No : 02531/84

Govt. of Tamil Nadu, Directorate of Medical and Rural Health Services CEA Permission No.: TNVLALL 2020000 8372 / 2019.
Tamil Nadu Pollution Control Board Permission No.: 19 BA, D 28501750/2019, 20GAD 28504164/2019

An ISO 2001:2015 Certified, ISAC Accreditation and
CMC Vellore Monthly Quality Control Lab. Or.No. : CMC/QC/2739/08, ISO No. : ISO915181388

18/34, Perumal Coil Street,
VALLIOOR - 627 117, Tirunelveli Dist.

☎ : 9487959009, ☎ : 04637 - 220798

Branch : Main Road, North Bazar,

✓ ANJUGRAMAM - 629 401, Kanyakumari Dist.

☎ : 8903966499 ☎ : 04652 - 267125



RESOLVING QC FAILURE

Steps taken to resolve out of range values for the month of November

- 1..Fresh reagent were prepared and samples were re-run
- 2.Reagent slots were reviewed
3. Sample storage and handling was reviewed
- 4.pipettes and instruments were checked



SOBHI DIAGNOSTIC CENTRE,
Main Road, North Bazar,
ANJUGRAMAM - 629 401,
Kanyakumari District.



We pray for your speedy recovery
MJF Ln. J. Sobitha Raj, M.A., B.Ed., M.A., D.M.L.T.,
Mrs. P. Sarojini Sobitha Raj, M.A., D.M.L.T., M.I.PH.&H

