

LAB MONTHLY SUMMARY



Lab Name **SOBHI DIAGNASTIC CENTER** Lab No **11625**
 Month **December** Year **2020**
 Constituent Group **Chemistry II**



Date of Result Entered : 14/12/2020

Date of Report Published : 31/12/2020

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Robonik (Semi automatic / Fully Automatic)	402	142.53	10.91	15.55	128 mg/dl	-0.93	1.55
2	UREA I	Urease UV / GLDH	Robonik (Semi automatic / Fully Automatic)	115	79.89	13.57	10.84	96.3 mg/dl	1.51	2.02
3	CREATININE I	Jaffes End point	Any Analyser	361	3.94	13.67	0.54	3.9 mg/dl	-0.07	0.06
4	T.BILIRUBIN I	OTHERS (Any Other Principles / Methods)	Any Analyser	163	2.18	21.31	0.46	3.9 mg/dl	3.74	0.07
5	T-PROTEIN I	Biuret - Colorimetric	Robonik (Semi automatic / Fully Automatic)	160	4.93	11.65	0.57	5.4 g/dl	0.82	0.09
6	ALBUMIN I	BCG - colorimetric	Robonik (Semi automatic / Fully Automatic)	142	2.97	11.04	0.33	2.7 g/dl	-0.82	0.06
7	URIC ACID I	OTHERS (Any Other Principles / Methods)	Any Analyser	94	4.22	21.12	0.89	5.1 mg/dl	0.99	0.18
8	CHOLESTEROL I	CHOD-PAP	Robonik (Semi automatic / Fully Automatic)	358	102.63	10.04	10.30	105.3 mg/dl	0.26	1.09
9	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Robonik (Semi automatic / Fully Automatic)	190	108.50	17.11	18.56	90.5 mg/dl	-0.97	2.69

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****



SOBHI Diagnostic Centre

A/c Computerised

Tamilnadu Govt Regd. No : 02531/84

18/34, Perumal Coil Street,
VALLIOOR - 627 117, Tirunelveli Dist.

* : 9487959009, ☎ : 04637 - 220798

Branch : Main Road, North Bazar,

✓ ANJUGRAMAM - 629 401, Kanyakumari Dist.

* : 8903966499 ☎ : 04652 - 267125

Govt. of Tamil Nadu, Directorate of Medical and Rural Health Services CEA Permission No.: TNVLALL 2020000 8372 / 2019.

Tamil Nadu Pollution Control Board Permission No.: 19 BA, D 28501750/2019, 20GAD 28504164/2019

An ISO 2001:2015 Certified, ISAC Accreditation and


CMC Vellore Monthly Quality Control Lab. Or.No. : CMC/QC/2739/08, ISO No. : ISO915181388




RESOLVING QC FAILURE

Steps taken to resolve out of range values for the month of December

1. Maintenance of instruments were done prior to sample analysis
2. Fresh reagent were prepared and samples were re-run
3. Sample storage and handling and instrument operation was reviewed
4. Internal quality control was done
5. Staff training was performed



SOBHI DIAGNOSTIC CENTRE,
Main Road, North Bazar,
ANJUGRAMAM - 629 401,
Kanyakumari District.



We pray for your speedy recovery
MJF Ln. J. Sobitha Raj, M.A., B.Ed., M.A., D.M.L.T.,
Mrs. P. Sarojini Sobitha Raj, M.A., D.M.L.T., M.I.P.H.&H

