

LAB MONTHLY SUMMARY



Lab Name **SOBHI DIAGNASTIC CENTER** Lab No **11625**  
 Month **January** Year **2021**  
 Constituent Group **Chemistry II**



Date of Result Entered : **17/01/2021**

Date of Report Published : **05/02/2021**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Robonik ( Semi automatic / Fully Automatic )	379	118.64	9.39	11.14	118 mg/dl	-0.06	1.14
2	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Robonik ( Semi automatic / Fully Automatic )	263	1.30	13.52	0.18	0.9 mg/dl	-2.27	0.02
3	T.BILIRUBIN II	Diazonium Salt ( Colorimetric ) / Jendrassik II	Robonik ( Semi automatic / Fully Automatic )	233	2.01	23.12	0.46	0.8 mg/dl	-2.61	0.06
4	T-PROTEIN II	Biuret - Colorimetric II	Robonik ( Semi automatic / Fully Automatic )	173	5.56	10.63	0.59	5.5 g/dl	-0.10	0.09
5	ALBUMIN II	BCG - Colorimetric II	Robonik ( Semi automatic / Fully Automatic )	167	3.33	11.35	0.38	2.6 g/dl	-1.93	0.06
6	URIC ACID II	Enzymatic / Uricase Colorimetric II	Robonik ( Semi automatic / Fully Automatic )	193	5.70	16.69	0.95	6.9 mg/dl	1.26	0.14
7	CHOLESTROL II	CHOD-PAP II	Robonik ( Semi automatic / Fully Automatic )	353	114.48	12.54	14.36	118.3 mg/dl	0.27	1.53

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)

Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
 CMC EQAS Co-Ordinator  
 Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*



# **SOBHI** *Diagnostic Centre*

A/c Computerised  
Tamilnadu Govt Regd. No : 02531/84

18/34, Perumal Coil Street,  
VALLIOOR - 627 117, Tirunelveli Dist.  
☎ : 9487959009, ☎ : 04637 - 220798



Branch : Main Road, North Bazar,  
✓ ANJUGRAMAM - 629 401, Kanyakumari Dist.  
☎ : 8903966499 ☎ : 04652 - 267125


Govt. of Tamil Nadu, Directorate of Medical and Rural Health Services CEA Permission No.: TNVLALL 2020000 8372 / 2019.  
Tamil Nadu Pollution Control Board Permission No.: 19 BA, D 28501750/2019, 20GAD 28504164/2019

An ISO 2001:2015 Certified, ISAC Accreditation and  
CMC Vellore Monthly Quality Control Lab. Dr.No. : CMC/QC/2739/08, ISO No. : ISO915181388

## RESOLVING QC FAILURE

Steps taken to resolve out of range values for the month of January

1. Maintenance of instruments were done prior to sample analysis
2. Fresh reagent were prepared and samples were re-run
3. Sample storage and handling and instrument operation was reviewed
4. Internal quality control was done
5. Staff training was performed



**SOBHI DIAGNOSTIC CENTRE,**  
Main Road, North Bazar,  
ANJUGRAMAM - 629 401,  
Kanyakumari District.



*We pray for your speedy recovery*  
MJF Ln. *J. Sobitha Raj*, M.A., B.Ed., M.A., D.M.L.T.,  
*Mrs. D. Sarojini Sobitha Raj*, M.A., D.M.L.T., M.I.P.H.&H

