

LAB MONTHLY SUMMARY



Lab Name **SOBHI DIAGNASTIC CENTER** Lab No **11625**
 Month **February** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : **24/02/2021**

Date of Report Published : **06/03/2021**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Robonik (Semi automatic / Fully Automatic)	397	242.04	11.04	26.71	227.2 mg/dl	-0.56	2.68
2	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Robonik (Semi automatic / Fully Automatic)	288	2.73	13.34	0.36	2.8 mg/dl	0.19	0.04
3	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Robonik (Semi automatic / Fully Automatic)	172	1.33	17.95	0.24	2.5 mg/dl	4.88	0.04
4	T-PROTEIN II	Biuret - Colorimetric II	Robonik (Semi automatic / Fully Automatic)	176	5.55	9.43	0.52	5.2 g/dl	-0.67	0.08
5	ALBUMIN II	BCG - Colorimetric II	Robonik (Semi automatic / Fully Automatic)	176	3.32	9.97	0.33	2.1 g/dl	-3.69	0.05
6	URIC ACID II	Enzymatic / Uricase Colorimetric II	Robonik (Semi automatic / Fully Automatic)	179	4.21	14.63	0.62	4.5 mg/dl	0.47	0.09
7	CHOLESTROL II	CHOD-PAP II	Robonik (Semi automatic / Fully Automatic)	382	115.49	10.63	12.27	113.5 mg/dl	-0.16	1.26

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****



SOBHI *Diagnostic Centre*

A/c Computerised
Tamilnadu Govt Regd. No : 02531/84

18/34, Perumal Coll Street,
VALLIOOR - 627 117, Tirunelveli Dist.
☎ : 9487959009, 📠 : 04637 - 220798



Branch : Main Road, North Bazar,
ANJUGRAMAM - 629 401, Kanyakumari Dist.
☎ : 8903966499, 📠 : 04652 - 267125

Govt. of Tamil Nadu, Directorate of Medical and Rural Health Services CEA Permission No.: TNVLALL 2020000 8372 / 2019.
Tamil Nadu Pollution Control Board Permission No.: 19 BA, D 28501750/2019, 20GAD 28504164/2019

An ISO 2001:2015 Certified, ISAC Accreditation and
CMC Vellore Monthly Quality Control Lab. Or.No. : CMC/QC/2739/08, ISO No. : ISO915181388

RESOLVING QC FAILURE

Steps taken to resolve out of range values for the month of February

- 1.Reagent slots were reviewed
- 2.Fresh reagent were prepared and samples were re-run
- 3.Sample storage and handling was reviewed
- 4.Pipettes and instruments were checked
- 5.Staff training was performed

SOBHI DIAGNOSTIC CENTRE,
Main Road, North Bazar,
ANJUGRAMAM - 629 401,
Kanyakumari District.



We pray for your speedy recovery
MJF Ln. J. Sobitha Raj, M.A., B.Ed., M.A., D.M.L.T.,
Mrs. P. Sarojini Sobitha Raj, M.A., D.M.L.T., MIPH & H

