



Urine Routine Analysis

ALL METHOD REPORT

Lab Code: 2083

Cycle - 10/2021
Round No - 1

Date: 12/02/2021

Parameters	Units	No.of. Participants	Consensus		Consensus Result	Your Value	Remark
			Number	%			
Specific Gravity	-	114	91	80%	1.010 - 1.020	1.020	Within consensus
pH	-	115	112	97%	6.0 - 7.0	7.0	Within consensus
Leukocytes	/µl	110	98	89%	1+ - 3+	3+	Within consensus
Nitrite	-	108	99	92%	Positive	Positive	Within consensus
Protein	mg/dl	115	102	89%	2+ - 4+	3+	Within consensus
Glucose	mg/dl	115	95	83%	3+ - 4+	3+	Within consensus
Ketones	-	115	97	84%	1+ - 3+	1+	Within consensus
Urobilinogen	mg/dl	115	98	85%	1+ - 3+	2+	Within consensus
Bilirubin	mg/dl	115	102	89%	3+	3+	Within consensus
Blood (Hemoglobin)	/µl	114	110	96%	3+ - 5+	3+	Within consensus

Legends	≥ 80% Consensus	(*) Outside Consensus	(.) Not Reported	(#)Late Result Submission
---------	-----------------	-----------------------	------------------	---------------------------

Chief Coordinator

Dr.Sanjay Mehrotra

Checked By:

Programme Director

Dr.Bandana Mehrotra



RML – Quality Assurance Program (RML – QAP)



RML - Quality Assurance Program

FINAL ASSESSMENT REPORT

JAN 2021 - DEC 2021

CYCLE - 10/2021

Provider

RML - QAP (Quality Assurance Program)
Under RML Research Foundation
B-171, Nirala Nagar,
Lucknow
India

Coordinator

Dr. Sanjay Mehrotra
Mr. Sohaib Farooqui

Contact us :-

Tel:- 0522-2788444, 2788666, 4034100,4077180

Email-Id:- rmlresearchfoundation@gmail.com
Website:- www.rmlpathology.com