


LAB MONTHLY SUMMARY

Lab Name **GURUNANAK DIAGNOSTICS**Lab No **9555**Month **March**Year **2021**Constituent Group **Chemistry I**Date of Result Entered : **19/03/2021**Date of Report Published : **01/04/2021**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	GOD-POD	Biosystems	230	169.20	6.59	11.15	170 mg/dl	0.07	1.47
2	UREA	Urease UV / GLDH	Biosystems	210	114.21	8.44	9.64	117.8 mg/dl	0.37	1.33
3	T-PROTEIN	Biuret - Colorimetric	Biosystems	230	5.33	9.02	0.48	5.6 g/dl	0.56	0.06
4	ALBUMIN	BCG - colorimetric	Biosystems	211	3.24	7.90	0.26	3.2 g/dl	-0.16	0.04
5	CALCIUM	Arsenazo III	Biosystems	138	10.92	12.80	1.40	4.8 mg/dl	-4.37	0.24
6	PHOSPHORUS	Molybdate UV / Phosphomolybdate complex	Biosystems	99	3.75	16.62	0.62	4 mg/dl	0.40	0.13
7	URIC ACID	Enzymatic / Uricase Colorimetric	Biosystems	219	6.77	9.32	0.63	7.28 mg/dl	0.81	0.09
8	CHOLESTEROL	CHOD-PAP	Biosystems	226	111.61	8.54	9.53	107.7 mg/dl	-0.41	1.27
9	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Biosystems	206	168.81	9.87	16.67	166.6 mg/dl	-0.13	2.32
10	AST	UV kinetic with PLP (P-5-P)	Any Analyser (Automation / Semi Automation)	586	161.74	12.69	20.53	182.7 U/L	1.02	1.70
11	ALT	UV kinetic with PLP (P-5-P)	Any Analyser (Automation / Semi Automation)	578	86.36	16.84	14.54	104.3 U/L	1.23	1.21
12	ALP	PNP AMP kinetic	Biosystems	163	217.81	12.01	26.17	222.7 U/L	0.19	4.10

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

******* End of Report *******