

# SRI SARAVANA ELECTRICAL SERVICE CENTRE

No. 4/9, Perumal Salai, New Perungalathur, Chennai - 600 063.

Specialist in : All kinds of Scientific Instruments and Laboratories Equipments,  
Blood Bank Equipments, B.O.D, Incubator, Ovens, Plasma Freezer, Auto Clave, Furnace, Microscopes,  
Laminar, Flow Unit, Colorimeter, Water Bath, Electrical balance and other Lab Equipments Sales & Services.

We are Undertaking Annual Maintenance Contract and Instruments Calibration.

Email : saravanascientific@yahoo.co.in

## @ CALIBRATION CERTIFICATE

CUSTOMER NAME AND ADDRESS  
THE BIOCLINICAL LABORATORY,  
(R.T.O. BUILDING)  
NO.4,SAI NAGAR MAIN ROAD,  
VIRUGAMBAKKAM,  
CHENNAI-600 092.

CALIBRATION CERTIFICATE NO : 3088/21  
DATE OF CALIBRATION : 09-02-2021  
NEXT CALIBRATION DATE : 09-02-2022

DESCRIPTION OF ITEM : PIPET  
TYPE : VARIABLE  
VOLUME : 5 µl.-120 µl.  
MAKE :  
SL.NO : 150083  
ROOM TEMP : 25°C

### RESULT OBSERVATIONS:

SL NO	MASTER PIPET	LAB PIPET	DIFF.IN.AC TUAL	REMARK
1	5 µl.	5.05 µl.	0.05 µl.	Accepted
	10 µl.	10.05 µl.	0.05 µl.	Accepted
	50 µl.	50.05 µl.	0.05 µl.	Accepted
	100 µl.	100.07 µl.	0.07 µl.	Accepted
	110 µl.	110.06 µl.	0.06 µl.	Accepted

- Remark: 1. CALIBRATION TRACEBLE TO NATIONAL STANDARDS  
2. OVER ALL DEVIATION IS WITH IN SPECIFIED ACCURACY HENCE THE INSTRUMENT IS OK.  
3. RESULTS ARE RELATED ONLY TO THE ITEM CALIBRATED  
4. MAXIMUM TOLERANCE PERMISSIBLE IS 0.1 µl.

AUTHORISED SIGNATORY & CALIBRATED BY.



*R. Durai Babu*  
R. DURAI BABU.  
SERVICE ENGINEER

# SRI SARAVANA ELECTRICAL SERVICE CENTRE

No. 4/9, Perumal Salai, New Perungalathur, Chennai - 600 063.

Specialist in : All kinds of Scientific Instruments and Laboratories Equipments,  
Blood Bank Equipments, B.O.D, Incubator, Ovens, Plasma Freezer, Auto Clave, Furnace, Microscopes,  
Laminar, Flow Unit, Colorimeter, Water Bath, Electrical balance and other Lab Equipments Sales & Services.

We are Undertaking Annual Maintenance Contract and Instruments Calibration.

Email : saravanascientific@yahoo.co.in

@

## CALIBRATION CERTIFICATE

CUSTOMER NAME AND ADDRESS  
THE BIOCLINICAL LABORATORY,  
(R.T.O. BUILDING)  
NO.4,SAI NAGAR MAIN ROAD,  
VIRUGAMBAKKAM,  
CHENNAI-600 092.

CALIBRATION CERTIFICATE NO : 3087/21  
DATE OF CALIBRATION : 09-02-2021  
NEXT CALIBRATION DATE : 09-02-2022

DESCRIPTION OF ITEM : PIPET  
SL.NO : 203045  
TYPE : VARIABLE  
VOLUME : 50 µl - 1000 µl.  
MAKE :  
ROOM TEMP : 25°C

**RESULT OF CALIBRATION OBSERVATIONS:**

SL NO	MASTERPIPET	LAB PIPET	DIFF.IN.ACTUAL	REMARK
1	50 µl.	50.05 µl.	0.05 µl.	Accepted
	100 µl.	100.05 µl.	0.05 µl.	Accepted
	200 µl.	200.08 µl.	0.08 µl.	Accepted
	500µl.	500.07 µl.	0.07 µl.	Accepted
	950 µl.	950.06 µl.	0.06 µl.	Accepted

- Remark: 1. CALIBRATION TRACEBLE TO NATIONAL STANDARDS  
2. OVER ALL DEVIATION IS WITH IN SPECIFIED ACCURACY HENCE THE INSTRUMENT IS OK.  
3. RESULTS ARE RELATED ONLY TO THE ITEM CALIBRATED  
4. MAXIMUM TOLERANCE PERMISSIBLE IS 0.1 µl.



AUTHORISED SIGNATORY & CALIBRATED BY.

*(Signature)*  
R. DURAI BABU.  
SERVICE ENGINEER