

SRI SARAVANA ELECTRICAL SERVICE CENTRE

4/9, PERUMAL SALAI, NEW PERUNGALATHUR. CHENNAI-600 063 No.

Specialist in: All kind of Scientific Instruments and Medical Equipments, Service Works Hospital, College, Clinical Lab, Blood Bank Equipments, Cooling Centrifuges, B.O.D, Incubator, Auto Clave, Furnace, Microscopes, Laminar Flow, Deep freezers, Incubators, Hot air ovens. And others Calorimeter,

We Are Undertaking Annual Maintenance Contract and Instruments Calibration.

@

CALIBRATION CERTIFICATE

CUSTOMER NAME AND ADDRESS
THE BIOCLINICAL LABORATORY,
(R.T.O. BUILDING)
NO.4, SAI NAGAR MAIN ROAD,
VIRUGAMBAKKAM,
CHENNAI-600 092.

CALIBRATION CERTIFICATE NO : 3071/21
DATE OF CALIBRATION : 09-02-2021
NEXT CALIBRATION DATE : 09-02-2022

DESCRIPTION OF ITEM : LABORATORY CENTRIFUGE. ELECTRA
MODEL NO : R.8C- /8/13
SL.NO : HGLC-10541
ROTOR HEAD : 16X15 ML . SOH
LOCATION : LABORATORY
MAKE : REMI

DETAILS OF EQUIPMENT USED FOR CALIBRATION:

SR.	DESCRIPTION	CERTIFIED BY	CERTIFICATE.NO.	VALIDITY
1.	DIGITAL TACHO METER. LUTRON MODEL. DT.2234.C. SL NO.L.207666	NORTH LAB CHENNAI. 4 NABL.	CERTIFICATE XEROX ENCLOSED	CERTIFICATE XEROX ENCLOSED

OBSERVATIONS:

SR.	INSTRUMENT SPEED DISPLAYED (RPM)	SPEED READING ON TACHOMETRE (RPM)	DIFF. IN. ACTUAL & TACHOMETRE (RPM)	REMARK
1.	1000	1060	60	Acceptable
2.	1750	1800	50	Acceptable
3.	2600	2650	50	Acceptable
4.	3000	3060	60	Acceptable
5.	4000	4050	50	Acceptable

Remark: 1. CALIBRATION TRACEBLE TO NATIONAL STANDARDS
2. OVERALL DEVIATION IS WITH IN SPECIFIED ACCURACY, HENCE THE INSTRUMENT IS OK.
3. RESULTS ARE RELATED ONLY TO THE ITEM CALIBRATED
4. MAXIMUM TOLERANCE PERMISSIBLE IS ± 100 RPM
5. TRACEABILITY XEROX COPY ENCLOSED

AUTHORIZED SIGNATURE & CALIBRATED BY

R. DURAI BABU
(Service Engineer)