

MK BEST CALIBRATION SERVICES



एमके बेस्ट केलिब्रेशन सर्विसेस

No. 27, F-2, 1st Floor, 2nd Street, Varalakshmi Nagar, Maduravoyal,
(Opp. MGR Engineering College), Chennai - 600 095.

Ph.: 044 - 23780211, Cell : 93802 66480 / 86958 18108 / 90032 77250

E-mail: mkbestcalibration@gmail.com, www.mkbestcalibrationservices.com



CERTIFICATE OF CALIBRATION



MQF/5.10/01

Certificate No	: MKBL-0083-003	Date of Issue	: 19.03.2021
Date of Calibration	: 19.03.2021	Recom.Due Date	: 18.03.2022
Customer Details :			
M/s: SHIFA HOSPITALS		SRF.No	: MKBL-21-03-0083
NO.82, KAILASAPURAM MIDDLE STREET,		Date of Receipt	: 19.03.2021
THIRUNELVELI - 627 011		Condition of UUC Received	: Satisfactory

Details Of Unit Under Calibration :			
Description	: PIPETTE	Identification No	: ----
Make	: BIO SYSTEM	Range	: 100 - 1000 µl
Model No	: ----	Least Count	: 1 µl
SL.No	: MC405297	Location	: ----

NABL TRACEABILITY DETAILS					
S.No	Description	Id.No/SI. No	Traceability	Certificate No	Validity
01	Electro Semi Micro Balance	MK/CAL-96/477904	National Standards	TVCSPL 20/06/415-01	02.06.2021

Cal. Procedure ref	EURAMET cg-19,OIML R40 & OIML R 43
Environmental Details :Temperature:	22.2 °C
Relative Humidity:	54 %

CALIBRATION RESULTS			
MASS & VOLUME CALIBRATION			
Sl.No	Standard Reading	UUC Reading	Error
	µl	µl	µl
1	100.0	100.17	0.17
2	200.0	200.21	0.21
3	500.0	500.28	0.28
4	700.0	700.31	0.31
5	1000.0	1000.42	0.42

Remarks :
1. The above UUC was Calibrated at Lab.

S. Chandra Bose
S.Chandra Bose
Calibration Engineer



L. Magesh
L.Magesh
Authorised Signatory

NABL ACCREDITED CALIBRATION LABORATORY AS PER ISO/IEC17025 : 2005

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION ISO 9001 : 2015.