

MK BEST CALIBRATION SERVICES



एमके बेस्ट केलिब्रेशन सर्विसेस

No. 27, F-2, 1st Floor, 2nd Street, Varalakshmi Nagar, Maduravoyal,
(Opp. MGR Engineering College), Chennai - 600 095.

Ph.: 044 - 23780211, Cell : 93802 66480 / 86958 18108 / 90032 77250

E-mail: mkbestcalibration@gmail.com, www.mkbestcalibrationservices.com



CERTIFICATE OF CALIBRATION



Certificate No	MKBL-0083-001	Date of Calibration	19.03.2021	Date of Issue	19.03.2021
SRF. No	MKBL-21-03-0083	Recom. Due Date	18.03.2022	Page No	1 of 1
CUSTOMER INFORMATION			UNIT UNDER CALIBRATION		
Name	M/s. SHIFA HOSPITALS	Description	REFRIGERATOR		
Address	NO.82, KAILASAPURAM MIDDLE STREET, THIRUNELVELI - 627 011.	Range	As below		
		Make	BLUE STAR		
		Model	----		
		Sub Zero No	----		
		Serial No	CHF500A17M20750		
		Location	----		
S.No	Description	Id.No/SI. No	Traceability	Certificate No	Validity
01	RTD Sensor	MK/CAL-52	National Standards	2020-21/CFC/MUM/03/17	14.07.2021
02	Handy Multi Function Calibrator	MK/CAL-142	National Standards	2020-21/CFC/MUM/03/2	27.05.2021
ENVIRONMENTAL CONDITIONS			REFERENCE STANDARD & ACCEPTANCE LIMIT		
Temperature	21.5°C		Uncertainty ± °C	1.2	
Humidity	57 % RH				
CALIBRATION RESULTS					
Parameter / Ranges	Standard Reading	Unit Under Calibration	Error Observed		
°C	°C	°C	°C		
2 - 8°C	4.00	4.16	-0.16		
Remarks :					
1. The Expanded Uncertainty associated with the results are calculated at a confidence level of approximately 95% with a Coverage factor of k=2.					
2. The above UUC was Calibrated at Lab.					
Calibrated by			Authorised Signatory		
 S.Chandra Bose Calibration Engineer			 L.Magesh Authorised Signatory		



NABL ACCREDITED CALIBRATION LABORATORY AS PER ISO/IEC17025 : 2005

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION ISO 9001 : 2015.