



MARS Medical SYSTEMS

Specialised Service For Medical Laboratory, Equipment & Diagnostic Reagents

L-236, Shastri Nagar, Delhi-110052

Ph. : 011-23659425, Mob. 9212015886, 8178430963

E-mail : marsmedicalsyst@yahoo.com

Ref. No.

Date05-09-2020.....

Calibration Certificate, Automated Hematology Analyzer

Name	NIRVANA DIAGNOSTIC CENTRE
Address	MEDICAL MORE, PGIMS ROAD, ROHTAK (HARYANA)-124001
Contact	+919729997769


Instrument identification				Calibration Passed	
Serial no.	373011	Date of Calibration	05-09-2020	Yes	Yes
Model	CELL-TECH PLUS	Calibrator Exp. Date	04-03-2021	Yes	Yes

Inspection LIST (Relevant information should be filled in)


	YES/NO
1. The Voltage of Instrument should be 220 V + 10%.	Yes
2. Ground Voltage should be less than 10 volts	Yes
3. No interference found near Instrument.	Yes
4. Reagents are ok.	Yes
5. Before Calibration instrument passed Background test.	Yes
6. Fresh sample/Control used.	Yes
7. Voltage of WBC and RBC chamber within range.	Yes

CALIBRATOR	Result	Target	CV %	REMARK
WBC	9.40	9.50	1.1%	OK
RBC	4.30	4.32	1.0%	OK
HGB	12.8	13.1	0.8%	OK
MCV	94.0	94.0	0.3%	OK
PLT	227	220	2.3%	OK

Calibration result is coming ok	✓	Remarks	Done
---------------------------------	---	---------	------


Service Engineer
05/09/2020

For Mars Medical Systems,


(Authorized Signatory)



MARS Medical Systems

Specialised Service For Medical Laboratory, Equipment & Diagnostic Reagents

L-236, Shastri Nagar, Delhi-110052
 Ph. : 011-23659425, Mob. 9212015886, 8178430963
 E-mail : marsmedicalsistem@yahoo.com

Ref. No.

Date 05-03-2021

Calibration Certificate, Automated Hematology Analyzer

Name	NIRVANA DIAGNOSTIC CENTRE
Address	MEDICAL MORE, PGIMS ROAD, ROHTAK (HARYANA)-124001
Contact	+919729997769


Instrument identification				Calibration Passed	
Serial no.	373011	Date of Calibration	05-03-2021	Yes	Yes
Model	CELL-TECH PLUS	Calibrator Exp Date	04-09-2021	Yes	Yes

Inspection LIST (Relevant information should be filled in)

	YES/NO
1. The Voltage of Instrument should be 220 V + 10%.	Yes
2. Ground Voltage should be less than 10 volts	Yes
3. No interference found near Instrument.	Yes
4. Reagents are ok.	Yes
5. Before Calibration instrument passed Background test.	Yes
6. Fresh sample/Control used.	Yes
7. Voltage of WBC and RBC chamber within range.	Yes

CALIBRATOR	Result	Target	CV %	REMARK
WBC	9.40	9.50	1.1%	OK
RBC	4.30	4.32	1.0%	OK
HGB	12.8	13.1	0.8%	OK
MCV	94.0	94.0	0.3%	OK
PLT	227	220	2.3%	OK

Calibration result is coming ok	✓	Remarks	Done
---------------------------------	---	---------	------


 Service Engineer
 05/03/2021

For Mars Medical Systems,

 (Authorized Signatory)