



MERAKI ENTERPRISES

Engineering services

meraki.719@gmail.com

+91-8809632355

Meenakshi complex, tower road Nawagunj ,Hazaribag -825301 | Hill view Road , Bariatu ,Ranchi -834009

End user: - Asarfi Hospital – Lab

Equipment	Micropipette
Make	Thermo
Sl.no	0W12641
Id no	As/lab/mp/1
Capacity	10 μ l
Date of calibration	17-09 -2020
Calibration due on	16-09-2021)
Environmental condition	(Temperature \pm 20) degree
Basis of calibration	Gravemetric method :distilled water
Traceability	Digital thermometer
	Balance Make :-KRSNA Model :- KI-AB-02G

Calibration Result

Sl.no	Duc set in μ l	Corresponding reading observed through ref.std.in Mg in balance	Tolerance \pm .05 mg
1.	10	10 .02 mg	

Remarks: The DUC has been calibrated over its above range .The reading observed are tabulated above

DUC: Device under Calibration

Physical status of the DUC :- OK

Engineer name & Sign :-Md salik Akhtar





MERAKI ENTERPRISES

Engineering services

meraki.719@gmail.com

+91-8809632355

Meenakshi complex, tower road Nawagunj ,Hazaribag -825301 | Hill view Road , Bariatu ,Ranchi -834009

End user: - Asarfi Hospital – Lab

Equipment	Micropipette
Make	Thermo
Sl.no	0W01447
Id no	As/lab/mp/2
Capacity	500µl
Date of calibration	17-09 -2020
Calibration due on	16-09-2021)
Environmental condition	(Temperature ± 20) degree
Basis of calibration	Gravemetric method :distilled water
Traceability	Digital thermometer
	Balance Make :-KRSNA Model :- KI-AB-02G

Calibration Result

Sl.no	Duc set in µl	Corresponding reading observed through ref.std.in Mg in balance	Tolerance ± .05 mg
1.	500	500 .04 mg	

Remarks: The DUC has been calibrated over its above range .The reading observed are tabulated above

DUC: Device under Calibration

Physical status of the DUC :- OK

Engineer name & Sign :-Md salik Akhtar

Meraki
Calibration
Md. Salik Akhtar



MERAKI ENTERPRISES

Engineering services

meraki.719@gmail.com

+91-8809632355

Meenakshi complex, tower road Nawagunj ,Hazaribag -825301 | Hill view Road , Bariatu ,Ranchi -834009

End user:-Asarfi Hospital-lab

Equipment	Micropipette
Make	Thermo
Sl.no	NW05577
Id no	As/lab/mp/3
Capacity	50 μ l
Date of calibration	17-09 -2020
Calibration due on	16-09-2021)
Environmental condition	(Temperature \pm 20) degree
Basis of calibration	Gravemetric method :distilled water
Traceability	Digital thermometer
	Balance Make :-KRSNA Model :- KI-AB-02G

Calibration Result

Sl.no	Duc set in μ l	Corresponding reading observed through ref.std.in Mg in balance	Tolerance \pm .05 mg
1.	50	50 .02 mg	

Remarks: The DUC has been calibrated over its above range .The reading observed are tabulated above

DUC: Device under Calibration

Physical status of the DUC :- OK

Engineer name & Sign :-Md salik Akhtar

Meraki
Calibration



MERAKI ENTERPRISES

Engineering services

meraki.719@gmail.com

+91-8809632355

Meenakshi complex, tower road Nawagunj ,Hazaribag -825301 | Hill view Road , Bariatu ,Ranchi -834009

End user: - Asarfi Hospital - Lab

Equipment	Micropipette
Make	Thermo
Sl.no	0W11753
Id no	As/lab/mp/5
Range	(02-20) μ l
Date of calibration	17-09 -2020
Calibration due on	16-09-2021)
Environmental condition	(Temperature \pm 20) degree
Basis of calibration	Gravemetric method :distilled water
Traceability	Digital thermometer
	Balance Make :-KRSNA Model :- KI-AB-02G

Calibration Result

Sl.no	Duc set in μ l	Corresponding reading observed through ref.std.in Mg in balance	Tolerance \pm .05 mg
1.	02	02.01 mg	
2.	10	10.03mg	
3.	20	20.02mg	

Remarks: The DUC has been calibrated over its above range .The reading observed are tabulated above

DUC: Device under Calibration

Physical status of the DUC :- OK

Engineer name & Sign :-Md salik Akhtar





MERAKI ENTERPRISES

Engineering services

meraki.719@gmail.com

+91-8809632355

Meenakshi complex, tower road Nawagunj ,Hazaribag -825301 | Hill view Road , Bariatu ,Ranchi -834009

End user: - Asarfi Hospital -Lab

Equipment	Micropipette
Make	Dragon lab
Sl.no	A0003817
Id no	As/lab/mp/7
Range	(500) μ l
Date of calibration	17-09 -2020
Calibration due on	16-09-2021)
Environmental condition	(Temperature \pm 20) degree
Basis of calibration	Gravemetric method :distilled water
Traceability	Digital thermometer
	Balance Make :-KRSNA Model :- KI-AB-02G

Calibration Result

Sl.no	Duc set in μ l	Corresponding reading observed through ref.std.in Mg in balance	Tolerance \pm .05 mg
1.	500	500.04 mg	

Remarks: The DUC has been calibrated over its above range .The reading observed are tabulated above

DUC: Device under Calibration

Physical status of the DUC :- OK

Engineer name & Sign :-Md salik Akhtar

Meraki
Calibration



MERAKI ENTERPRISES

Engineering services

meraki.719@gmail.com

+91-8809632355

Meenakshi complex, tower road Nawagunj ,Hazaribag -825301 | Hill view Road , Bariatu ,Ranchi - 834009

End user: - Asarfi Hospital -Lab

Equipment	Micropipette
Make	Dragon lab
Sl.no	K0006986
Id no	As/lab/mp/8
Range	(1000) μ l
Date of calibration	17-09 -2020
Calibration due on	16-09-2021)
Environmental condition	(Temperature \pm 20) degree
Basis of calibration	Gravemetric method :distilled water
Traceability	Digital thermometer
	Balance Make :-KRSNA Model :- KI-AB-02G

Calibration Result

Sl.no	Duc set in μ l	Corresponding reading observed through ref.std.in Mg in balance	Tolerance \pm .05 mg
1.	1000	1000.04 mg	

Remarks: The DUC has been calibrated over its above range .The reading observed are tabulated above

DUC: Device under Calibration

Physical status of the DUC :- OK

Engineer name & Sign :-Md salik Akhtar





MERAKI ENTERPRISES

Engineering services

meraki.719@gmail.com

+91-8809632355

Meenakshi complex, tower road Nawagunj ,Hazaribag -825301 | Hill view Road , Bariatu ,Ranchi - 834009

End user: - Asarfi Hospital -Lab

Equipment	Micropipette
Make	ERBA
Sl.no	011615692
Id no	As/lab/mp/9
Range	(100-1000) μ l
Date of calibration	17-09 -2020
Calibration due on	16-09-2021)
Environmental condition	(Temperature \pm 20) degree
Basis of calibration	Gravemetric method :distilled water
Traceability	Digital thermometer
	Balance Make :-KRSNA Model :- KI-AB-02G

Calibration Result

Sl.no	Duc set in μ l	Corresponding reading observed through ref.std.in Mg in balance	Tolerance \pm .05 mg
1.	100	100.04 mg	
2	500	500.03mg	
3.	1000	1000.02 mg	

Remarks: The DUC has been calibrated over its above range .The reading observed are tabulated above

DUC: Device under Calibration

Physical status of the DUC :- OK

Engineer name & Sign :-Md salik Akhtar





MERAKI ENTERPRISES

Engineering services

meraki.719@gmail.com

+91-8809632355

Meenakshi complex, tower road Nawagunj ,Hazaribag -825301 | Hill view Road , Bariatu ,Ranchi -834009

End user: -Asarfi Hospital-lab

Equipment	Micropipette
Make	Thermo
Sl.no	OW14171
Id no	As/lab/mp/4
Range	(20-200) μ l
Date of calibration	17-09 -2020
Calibration due on	16-09-2021)
Environmental condition	(Temperature \pm 20) degree
Basis of calibration	Gravemetric method :distilled water
Traceability	Digital thermometer
	Balance Make :-KRSNA Model :- KI-AB-02G

Calibration Result

Sl.no	Duc set in μ l	Corresponding reading observed through ref.std.in Mg in balance	Tolerance \pm .05 mg
1.	20	20.02 mg	
2.	100	100.03mg	
3.	200	200.04mg	

Remarks: The DUC has been calibrated over its above range .The reading observed are tabulated above

DUC: Device under Calibration

Physical status of the DUC :- OK

Engineer name & Sign :-Md salik Akhtar





MERAKI ENTERPRISES

Engineering services

meraki.719@gmail.com

+91-880963235

Meenakshi complex, tower road Nawagunj ,Hazaribag -825301 | Hill view Road , Bariatu ,Ranchi -834009

End user: - Asarfi Hospital -Lab

Equipment	Micropipette
Make	Dragon lab
Sl.no	M0060640
Id no	As/lab/mp/6
Range	(100-1000) μ l
Date of calibration	17-09 -2020
Calibration due on	16-09-2021)
Environmental condition	(Temperature \pm 20) degree
Basis of calibration	Gravemetric method :distilled water
Traceability	Digital thermometer
	Balance Make :-KRSNA Model :- KI-AB-02G

Calibration Result

Sl.no	Duc set in μ l	Corresponding reading observed through ref.std.in Mg in balance	Tolerance \pm .05 mg
1.	100	100.01 mg	
2.	500	500.03mg	
3.	1000	1000.02mg	

Remarks: The DUC has been calibrated over its above range .The reading observed are tabulated above

DUC: Device under Calibration

Physical status of the DUC :- OK

Engineer name & Sign :-Md salik Akhtar



