



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments
304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092
Phone / Fax : (011) 22453259, 22434362, 9599482307
E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

TEST REPORT

Page-1 of 2

TEST REPORT OF BIO SAFETY CABINET

Certificate No:	20085000074	Issued On:	31/08/2020
Job Identification No:	ICS/C/08-5/74	Calibration Date:	29/08/2020
Ref. No:	SRF- Dated: 27/08/2020	Next Calibration Date:	28/08/2021

CALIBRATED FOR: M/s Parmar Hospital,
Bela Road, Ropar(Punjab).

EQUIPMENT DESCRIPTION

Name	Bio Safety Cabinet, Equipped with HEPA Filter, Size: 600 X 600 mm		
Make	An Biotech	Model	---
Serial No.	---	I.D. No.	---
Location	---	Testing Site	On-Site

TEST REQUIRED

Sn	Name Of The Test	Specified Classification	Acceptance Criteria	Test Method
1.	HEPA Filter Integrity Efficiency Test	H-14 as per BS EN 1822-1:1998	≤ 0.025% Penetration	ISO 14644-3
2.	Air Borne Particle Measurement	Cleanliness Class ISO 5 as per ISO 14644-1	< 3520/m ³ for 0.5 μm	ISO 14644-3
3.	Air Velocity Measurement	As per EN 12469	0.25 m/s to 0.50 m/s	EN 12469

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 3.0° C	Humidity	50 ± 10 % RH
-------------	------------------	----------	--------------

STANDARD EQUIPMENT DETAILS Traceable to National Standards and NIST, U.S.A

Name	Make/Model	Certificate No.	Calibration Agency	Calibration Date	Valid Upto
DOP Touch Photometer	DOP, UK/ DOP touch 3000	MKC/017/2020- 21	MK Tech., Delhi	10/06/2020	10/06/2021
Air Particle Counter	CLIMET/ CI-752x	MKC/018/2020- 21	MK Tech., Delhi	10/06/2020	10/06/2021
Thermal Anemometer	TESTO/ 405-V1	CAW 127 2006 007	FCRI, Palakkad	15/06/2020	15/06/2021

Other Equipment Used: Aerosol Generator & Thermo-Hygrometer

Procedure No.: ICS/T/SOP-01

- Note :**
- This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
 - The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
 - Total liability of our Organisation is limited to the invoiced amount.
 - Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
 - In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
 - The calibration certificate/Test Report is valid only for the condition of the UVC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by

RAJESH KAPOOR
Authorised Signatory