



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments
304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307

E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF BLOOD MIXER

Certificate No:	20900001732	Issued On:	31/08/2020
ULR No:	CC214620900001732F	Calibration Date:	29/08/2020
Job Identification No:	ICS/C/AS/08/1732	Next Calibration Date:	28/08/2021
Ref. No:	SRF, Dated -27/08/2020		

CALIBRATED FOR: M/s Gurunanak Diagnostics,
Parmar Hospital, Bela Road, Ropar(Punjab).

EQUIPMENT DESCRIPTION

Name	RPM Measurement for Blood Mixer		
Make/Model	Labline	Visual Inspection	Ok
Serial No.	6469	I.D. No.	BM-01
Range	At 30 RPM	Least Count	---
Location	---	Calibration Site	On-Site

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 2.0° C	Humidity	45 to 75 % RH
-------------	------------------	----------	---------------

STANDARD REFERENCE DETAILS

Traceable to National Standard (NPL, New Delhi)

Name of the Standard	Make/Model	Certificate No.	Calibration Agency	Date of Calibration	Valid up to
Digital Tachometer	Lutron/ DT-2234C	V-200320-1-1	Godrej, Gujrat	25/05/2020	25/05/2021

USED EQUIPMENT DETAILS: Not Applicable

PRINCIPLE/METHODOLOGY OF CALIBRATION: As per Calibration Procedure No.: ICS/CAL/SOP-RPM01
(Comparison Method)

RESULTS: Mechanical Calibration
MEASUREMENT: (in RPM)

U.U.C. SPEED KNOB' POSITION	CORRECTED VALUE OF TACHOMETER READING	Uncertainty at approx 95% C.L and coverage factor k=2
Reading # 1	29.2	± 2
Reading # 2	29.5	
Reading # 3	30.1	
Reading # 4	29.7	
Reading # 5	30.2	

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

END OF REPORT

Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.

2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.

3. Total liability of our Organisation is limited to the invoiced amount.

4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.

5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.

6. The calibration certificate/Test Report is valid only for the condition of the UVC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by

RAJESH KAPOOR
Authorised Signatory