

# Indian Calibration Services



An Exclusive House for Validation & Calibration of Analytical Instruments  
 304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092

Phone / Fax : (011) 22453259, 22434362, 9599482307  
 E-mail : indiancalibration@indiancalibration.com Website : www.multitechs.com

## CALIBRATION CERTIFICATE


Page 1 of 2

### CALIBRATION CERTIFICATE OF REFRIGERATOR

Certificate No:	20100008396	Issued On:	31/08/2020
ULR No:	CC214620100008396F	Calibration Date:	29/08/2020
Job Identification No:	ICS/C/T/08/8396	Next Calibration Date:	28/08/2021
Ref. No:	SRF, Dated -27/08/2020	M/s Gurnanak Diagnostics, Pamar Hospital, Bela Road, Ropar(Punjab)	

EQUIPMENT DESCRIPTION			
Name	Refrigerator	Visual Inspection	OK
Make/Model	LG	I.D. No.	RF-01
Serial No.	---	Least Count	---
Range	2 to 8°C	Calibration Site	On-Site
Location	---	Humidity	50 ± 10 % RH
Temperature	25.0 °C ± 3.0 °C	STANDARD REFERENCE DETAILS	
Traceable to National Standard (NPL, New Delhi)			
Name of the Standard	Make/Model	Certificate No.	Calibration Agency/Traceability
Temp. & Hum. (RH) Data Logger	HTC/ICS/T/INS/17-25	2010000002-10	Indian Calibration Services, Delhi
USED EQUIPMENT DETAILS	Calibrated RTD Sensors	As per Calibration Procedure No.: ICS-CAL/SOP-TH02, ITS-90 (Measurement Method)	
PRINCIPLE/METHODOLOGY OF CALIBRATION:			
RESULTS: Thermal Calibration All readings are in "°C"			

Set Value in °C	S-1	S-2	S-3	S-4	S-5	S-6	S-7	S-8	S-9	Uncertainty(with 95% C.L. at k=2) (°C)
2 to 8	3.3	3.4	3.8	5.0	5.1	4.8	4.6	4.7	4.6	± 1.83
Max Stability:-	0.4									
Max Uniformity:-	1.3									



Refrigerator

Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.  
 2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.  
 3. Total liability of our Organisation is limited to the invoiced amount.  
 4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.  
 5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.  
 6. The calibration certificate/Test Report is valid only for the condition of the UVC at the time under stated condition of calibration.

Checked by: *[Signature]*  
 Authorised Signatory: *[Signature]*  
 RAJESH KAPOOR  
 Indian Calibration Services

For and on behalf of