Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments
304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092
Phone / Fax : (011) 22453259, 22434362, 9599482307
E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

NAME: REFRIGERATOR Inst. I.D.: RF-01

Calibration Date: 29/08/2020

		S-1	S-2	S-3	S-4	S-5	S-6	S-7	S-8	S-9
DATE	TIME	°C	°C	°C	°C	°C	°C	°C	°C	Center °C
29-08-2020	10:45:13	3.2	3.3	3.4	4.9	5.0	4.8	4.5	4.6	4.5
29-08-2020	10:46:13	3.3	3.4	3.5	5.0	5.2	4.9	4.7	4.8	4.7
29-08-2020	10:47:13	3.3	3.4	3.5	5.0	5.1	4.8	4.6	4.7	4.6
29-08-2020	10:48:13	3.2	3.3	3.4	4.9	5.0	4.8	4.5	4.6	4.5
29-08-2020	10:49:13	3.3	3.4	3.5	5.1	5.2	4.9	4.7	4.8	4.7
29-08-2020	10:50:13	3.3	3.4	3.5	5.0	5.1	4.8	4.6	4.7	4.6
29-08-2020	10:51:13	3.2	3.3	3.4	4.9	5.1	4.8	4.5	4.6	4.5
29-08-2020	10:52:13	3.4	3.5	3.6	5.1	5.2	4.9	4.7	4.8	4.7
29-08-2020	10:53:13	3.3	3.4	3.5	5.0	5.1	4.8	4.6	4.7	4.6
29-08-2020	10:54:13	3.2	3.3	3.4	4.9	5.0	4.7	4.5	4.6	4.5
29-08-2020	10:55:13	3.2	3.3	3.4	4.8	4.9	4.7	4.4	4.5	4.4
29-08-2020	10:56:13	3.1	3.2	3.3	4.7	4.8	4.6	4.3	4.4	4.3
29-08-2020	10:57:13	3.1	3.2	3.3	4.7	4.8	4.5	4.3	4.4	4.3
29-08-2020	10:58:13	3.0	3.1	3.2	4.6	4.7	4.5	4.2	4.3	4.2
29-08-2020	10:59:13	3.0	3.1	3.2	4.6	4.7	4.4	4.2	4.3	4.2
29-08-2020	11:00:13	3.2	3.3	3.4	4.9	5.0	4.7	4.5	4.6	4.5
29-08-2020	11:01:13	3.2	3.3	3.4	4.9	5.0	4.7	4.5	4.6	4.5
29-08-2020	11:02:13	3.2	3.3	3.4	4.9	5.0	4.8	4.5	4.6	4.5
29-08-2020	11:03:13	3.3	3.3	3.4	4.9	5.1	4.8	4.5	4.6	4.5
29-08-2020	11:04:13	3.3	3.4	3.5	4.9	5.1	4.8	4.5	4.6	4.5
29-08-2020	11:05:13	3.3	3.4	3.5	5.0	5.1	4.8	4.6	4.6	4.6
29-08-2020	11:06:13	3.3	3.4	3.5	5.0	5.1	4.9	4.6	4.7	4.6
29-08-2020	11:07:13	3.3	3.4	3.6	5.1	5.2	4.9	4.7	4.8	4.7
29-08-2020	11:08:13	3.4	3.5	3.6	5.1	5.3	5.0	4.7	4.8	4.7
29-08-2020	11:09:13	3.4	3.5	3.6	5.2	5.3	5.0	4.8	4.9	4.8
29-08-2020	11:10:13	3.5	3.6	3.7	5.3	5.4	5.1	4.8	4.9	4.8
29-08-2020	11:11:13	3.4	3.5	3.7	5.2	5.3	5.0	4.7	4.8	4.7
29-08-2020	11:12:13	3.3	3.4	3.8	5.1	5.2	4.9	4.7	4.7	4.7
29-08-2020	11:13:13	3.4	3.5	3.7	5.1	5.3	5.0	4.7	4.8	4.7
29-08-2020	11:14:13	3.3	3.4	3.8	5.0	5.2	4.9	4.6	4.7	4.6
29-08-2020	11:15:13	3.3	3.4	3.8	4.9	5.1	4.8	4.5	4.6	4.5
Avg.		3.3	3.4	3.8	5.0	5.1	4.8	4.6	4.7	4.6
Stability		0.2	0.2	0.3	0.3	0.4	0.3	0.3	0.3	0.3
Max.stab.		0.4								
Uniformity		1.3	1.2	0.7	0.4	0.5	0.2	0.0	0.1	0.0
Max. Unif.		1.3								

REMARKS:

- The recommended date for next calibration is mentioned, as asked by the customer.

END OF REPORT

Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
3. Total liability of our Organisation is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
6. The calibration certificate/Test Report is valid only for the condition of the UVC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by

RAJESH KAPOOR
Authorised Signatory
Operation Head