

Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments
 304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092
 Phone / Fax : (011) 22453259, 22434362, 9599482307
 E-mail : indiancalibration@indiancalibration.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

CALIBRATION CERTIFICATE OF MICROSCOPE

Certificate No:	20085000073	Issued On:	31/08/2020
Job Identification No:	ICS/C/08-5/73	Calibration Date:	29/08/2020
Ref. No:	SRF - Dated: 27/08/2020	Next Calibration Date:	28/08/2021
CALIBRATED FOR:		M/s Gurnanak Diagnostics, Parmar Hospital, Bela Road, Ropar(Punjab).	

EQUIPMENT DESCRIPTION			
Name	Microscope	Visual Inspection	OK
Make/Model	Olympus	I.D. No.	MS-01
Serial No.	---	Eye Piece/Least Count	---
Range	4X, 10X	Calibration Site	On Site

ENVIRONMENTAL CONDITIONS			
Temperature	20.0 °C ± 2.0 °C	Humidity	50 ± 10 % RH

STANDARD REFERENCE DETAILS			
Traceable to National Standard			
Name of the Standard	Make/Model	Certificate No.	Calibration Agency/Traceability
Slip Gauges Set	Aditya/ 2938	TI/F/SGS/061/20	Tanson Instrument, Faridabad
Glass Scale	ICS/STD/91	TI/IGS/005/18	Tanson, Fbd
USED EQUIPMENT DETAILS	As per Calibration Procedure No.: ICS/CAL/SOP-D31		
PRINCIPLE/METHODOLOGY OF CALIBRATION:	N/A		

RESULTS:			
Parameter	Specified Magnification	STD Value	Measured Value On Magnification (mm)
Magnification	4 X	1 mm	4.0
Magnification	10 X	1 mm	10.0
Uncertainty at approx 95% C.L and coverage factor k=2		± 0.18% of magnification	

REMARKS:
 • The recommended date for next calibration is mentioned, as asked by the customer.
 *** END OF REPORT ***

Note: 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
 2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
 3. Total liability of our Organisation is limited to the invoiced amount.
 4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
 5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
 6. The calibration certificate/Test Report is valid only for the condition of the UVC at the time under stated condition of calibration.

Checked by: *[Signature]*
 Authorised Signatory: *[Signature]*
 For and on behalf of Indian Calibration Services
 Operation Head: *[Signature]*