

Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments
 304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092
 Phone / Fax : (011) 22453259, 22434362, 9599482307
 E-mail : indiancalibrationservices@gmail.com Website : www.multitechs.com



CALIBRATION CERTIFICATE OF DIGITAL THERMOMETER WITH SENSOR

Certificate No:	20100008303	Issued On:	29/08/2020
ULR No:	CC214620100008303F	Calibration Date:	27/08/2020
Job Identification No:	ICS/C/T/08/8303	Next Calibration Date:	26/08/2021
Ref. No:	SRF, Dated -25/08/2020		

CALIBRATED FOR:

M/s Gurnanak Diagnostics,
 Parmar Hospital, Bela Road, Ropar (Punjab)

EQUIPMENT DESCRIPTION

Name	Digital Thermometer with Sensor
Make/Model	---
Serial No.	---
Range	-50 to 110°C
Location	---
Visual Inspection	OK
I.D. No.	DTM-01
Least Count	0.1°C
Calibration Site	In-Lab

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 3.0 °C	Humidity	50 ± 10 % RH
-------------	------------------	----------	--------------

STANDARD REFERENCE DETAILS

Name of the Standard	Make/Model	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
RTD Sensor with Digital Indicator	Udian	ERTL(N)/2020/AR0878	ERTL (NORTH)	25/02/2020	25/02/2021

USED EQUIPMENT DETAILS

Methanol Bath, Water Bath & Dry Block Calibrator	
As per Calibration Procedure No.: ICS-CAL/SOP-TH02, ITS-90 (Comparison Method)	

RESULTS: Thermal Calibration (On Working Range)

U.U.C. READING (in °C)	CORRECTED VALUE OF STANDARD (in °C) (Avg.)	Uncertainty at approx 95% C.L and coverage factor k=2 (in °C)
-20.0	-19.925	± 0.18
30.0	30.051	± 0.28
50.0	50.097	± 0.28
80.0	80.156	± 0.28
100.0	100.229	± 0.41

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.
 *** END OF REPORT ***

Note: 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
 2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
 3. Total liability of our Organisation is limited to the invoiced amount.
 4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
 5. In case any reconfirmation of contents of this Calibration Certificate is required, Please contact our office.
 6. The calibration certificate/Test Report is valid only for the condition of the UVC at the time under stated condition of calibration.

Checked by: *[Signature]*

For and on behalf of: **Indian Calibration Services**
[Signature]
 RAJESH KAPOOR
 Authorised Signatory
 (Person Here)