


LAB MONTHLY SUMMARY

Lab Name **GURUNANAK DIAGNOSTICS**Lab No **9555**Month **February**Year **2021**Constituent Group **Chemistry I**Date of Result Entered : **19/02/2021**Date of Report Published : **06/03/2021**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	GOD-POD	Biosystems	220	258.67	6.30	16.30	209.2 mg/dl	-3.03	2.20
2	UREA	Urease UV / GLDH	Biosystems	211	43.63	10.03	4.38	37.3 mg/dl	-1.45	0.60
3	T-PROTEIN	Biuret - Colorimetric	Biosystems	218	5.18	6.41	0.33	3.9 g/dl	-3.86	0.04
4	ALBUMIN	BCG - colorimetric	Biosystems	207	3.20	8.60	0.28	2.49 g/dl	-2.58	0.04
5	CALCIUM	Arsenazo III	Biosystems	133	10.22	8.97	0.92	3 mg/dl	-7.85	0.16
6	PHOSPHORUS	Molybdate UV / Phosphomolybdate complex	Biosystems	96	5.43	12.88	0.70	7 mg/dl	2.24	0.14
7	URIC ACID	Enzymatic / Uricase Colorimetric	Biosystems	202	8.79	8.10	0.71	8 mg/dl	-1.11	0.10
8	CHOLESTEROL	CHOD-PAP	Biosystems	208	107.68	7.40	7.97	70.8 mg/dl	-4.63	1.11
9	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Biosystems	207	94.25	11.32	10.66	79.2 mg/dl	-1.41	1.48
10	AST	UV kinetic with PLP (P-5-P)	Any Analyser (Automation / Semi Automation)	559	84.74	14.08	11.93	73.2 U/L	-0.97	1.01
11	ALT	UV kinetic with PLP (P-5-P)	Any Analyser (Automation / Semi Automation)	444	96.68	15.88	15.35	86.9 U/L	-0.64	1.46
12	ALP	PNP AMP kinetic	Biosystems	170	94.88	10.32	9.79	108.9 U/L	1.43	1.50

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

******* End of Report *******