



Indian Calibration Services

An Exclusive House for Validation & Calibration of Analytical Instruments
304 & 313, Laxmi Deep, District Centre, Near Nirman Vihar Metro Station, Laxmi Nagar, Delhi-110092
Phone / Fax : (011) 22453259, 22434362, 9599482307
E-mail : indiancalibrationservices@gmail.com Website : www.multitechics.com

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE OF GLASS THERMOMETER

Certificate No:	20100008397	Issued On:	31/08/2020
ULR No:	CC214620100008397F	Calibration Date:	29/08/2020
Job Identification No:	ICS/C/T/08/8397	Next Calibration Date:	28/08/2021
Ref. No:	SRF, Dated -27/08/2020		

CALIBRATED FOR: M/s Gurunanak Diagnostics,
Parmar Hospital, Bela Road, Ropar(Punjab).

EQUIPMENT DESCRIPTION

Name	Glass Thermometer		
Make/Model	---	Visual Inspection	Ok
Serial No.	---	I.D. No.	INC-01
Range	0°C to 110°C	Least Count	1°C
Location	Incubator	Calibration Site	On-Site

ENVIRONMENTAL CONDITIONS

Temperature	25.0 °C ± 3.0° C	Humidity	50 ± 10 % RH
-------------	------------------	----------	--------------

STANDARD REFERENCE DETAILS

Traceable to National Standard (NPL, New Delhi)

Name of the Standard	Make/Model	Certificate No.	Calibration Agency/Traceability	Date of Calibration	Valid up to
Digital Temp. Indicator with Sensor	"Lutron"/ ICS/T/INS/41	20100000032	Indian Calibration Services, Delhi	02/01/2020	02/01/2021

USED EQUIPMENT DETAILS	Water Bath
PRINCIPLE/METHODOLOGY OF CALIBRATION:	As per Calibration Procedure No.: ICS/CAL/SOP-TH01 (ITS-90) (Comparison Method)

RESULTS: Thermal Calibration
All readings are in "° C"

U.U.C. READING	CORRECTED VALUE OF STANDARD	Uncertainty at approx 95% C.L and coverage factor k=2
30	30.06	± 0.65
35	35.13	± 0.65
37	37.20	± 0.65
45	45.29	± 0.65
60	60.41	± 0.71

REMARKS: The recommended date for next calibration is mentioned, as asked by the customer.

END OF REPORT

- Note : 1. This report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.
2. The result listed refer only to the tested samples and applicable parameters. Endorsement of products is neither inferred nor implied.
3. Total liability of our Organisation is limited to the invoiced amount.
4. Samples will be destroyed after one month from the date of issue of Calibration Certificate unless otherwise specified.
5. In case any reconfirmation of contents of this Calibration Certificate is required. Please contact our office.
6. The calibration certificate/Test Report is valid only for the condition of the UVC at the time under stated condition of calibration.

For and on behalf of
Indian Calibration Services

Checked by

RAJESH KAPOOR
Authorised Signatory