

HIN/MED/2021-22/040210

04th April, 2021

CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer of **HORIBA MEDICAL: ABX Yumizen H550** bearing serial no: - **012YAXH03186** installed at **St Theresa Hospital, Hyderabad** was calibrated with calibrator on **04th April, 2021**.

Calibrator : **ABX MINOCAL**
Lot No : **CX457**
Expiry Date : **5th May, 2021**

The reports of Blank cycle, Repeatability and Calibration values were all found in acceptable range.

Next Calibration is due on **03rd April, 2022**.



M.OMPRASKASH

ENGINEER –CS

For **Horiba India Pvt Ltd**