

LAB MONTHLY SUMMARY



Lab Name **Priya Clinical Laboratory** Lab No **11642**
 Month **March** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 08/03/2021

Date of Report Published : 01/04/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	791	160.91	10.47	16.84	138.5 mg/dl	-1.33	1.20
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	914	43.19	14.15	6.11	39.5 mg/dl	-0.60	0.40
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	491	1.67	11.16	0.19	1.9 mg/dl	1.24	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	658	3.60	13.31	0.48	2.6 mg/dl	-2.09	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	595	5.89	9.22	0.54	7.1 g/dl	2.23	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	599	3.52	9.28	0.33	4.3 g/dl	2.39	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	799	7.29	15.23	1.11	5.1 mg/dl	-1.97	0.08
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	717	118.97	12.88	15.32	108.4 mg/dl	-0.69	1.14
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	665	105.23	19.47	20.49	105.5 mg/dl	0.01	1.59

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email:clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****

