

LAB MONTHLY SUMMARY



Lab Name **Priya Clinical Laboratory** Lab No **11642**
 Month **February** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 12/02/2021

Date of Report Published : 06/03/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	747	251.19	10.64	26.72	215.7 mg/dl	-1.33	1.95
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	856	53.59	14.85	7.96	56.3 mg/dl	0.34	0.54
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	452	2.80	10.81	0.30	3.3 mg/dl	1.66	0.03
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	532	1.22	19.52	0.24	0.5 mg/dl	-3	0.02
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	580	5.56	10.15	0.56	5.1 g/dl	-0.82	0.05
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	570	3.31	9.68	0.32	4 g/dl	2.16	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	582	3.94	17.52	0.69	4.2 mg/dl	0.38	0.06
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	655	110.72	11.41	12.63	125.9 mg/dl	1.20	0.99
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	504	90.04	16.84	15.16	89.1 mg/dl	-0.06	1.35

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email:clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss

CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****

