

LAB MONTHLY SUMMARY



Lab Name **Priya Clinical Laboratory** Lab No **11642**
 Month **January** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 11/01/2021

Date of Report Published : 05/02/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	695	119.49	9.92	11.85	135.7 mg/dl	1.37	0.90
2	UREA II	Others (Urease Berthelot / Ned Dye) II	Any Analyser (Automation / Semi Automation)	806	33.34	17.16	5.72	31.5 mg/dl	-0.32	0.40
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Transasia / Erba	428	1.37	12.84	0.18	1.5 mg/dl	0.74	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	596	1.93	20.18	0.39	1 mg/dl	-2.39	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	527	5.66	9.56	0.54	5.1 g/dl	-1.04	0.05
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	527	3.38	9.61	0.32	3.1 g/dl	-0.86	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	696	5.57	16.35	0.91	4.6 mg/dl	-1.06	0.07
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	637	111.91	12.83	14.35	119.4 mg/dl	0.52	1.14
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	577	129.39	15.04	19.46	126.1 mg/dl	-0.17	1.62

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email:clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator

Christian Medical College, Vellore

***** End of Report *****

