

## LAB MONTHLY SUMMARY

Lab Name **Priya Clinical Laboratory** Lab No **11642**Month **November** Year **2020**Constituent Group **Chemistry II**Date of Result Entered : **23/11/2020**Date of Report Published : **05/12/2020**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE I	GOD-POD	Transasia / Erba	590	392.81	10.10	39.66	378.9 mg/dl	-0.35	3.27
2	UREA I	Ned Dye	Any Analyser	77	58.62	10.65	6.24	58.1 mg/dl	-0.08	1.42
3	CREATININE I	Jaffes Kinetic - Alkaline picrate	Transasia / Erba	498	7.97	11.62	0.93	7.8 mg/dl	-0.18	0.08
4	T.BILIRUBIN I	Diazonium salt ( Colorimetric ) / Jendrassik	Transasia / Erba	495	1.69	18.36	0.31	0.9 mg/dl	-2.55	0.03
5	T-PROTEIN I	Biuret - Colorimetric	Transasia / Erba	424	5.34	9.16	0.49	8.4 g/dl	6.24	0.05
6	ALBUMIN I	BCG - colorimetric	Transasia / Erba	462	3.23	9.21	0.30	3.2 g/dl	-0.10	0.03
7	CHOLESTEROL I	CHOD-PAP	Transasia / Erba	548	110.43	11.37	12.56	163.2 mg/dl	4.2	1.07
8	TRIGLYCERIDE I	GPO-PAP / Enzymatic Colorimetric / End Point	Transasia / Erba	525	148.27	15.04	22.30	194.1 mg/dl	2.06	1.95

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between $\pm 1.0$ to $\pm 2.0$	Good.
Between $\pm 2.0$ to $\pm 3.0$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)

Contact Number: 0416-2283102

**Dr. Pamela Christudoss**  
**CMC EQAS Co-Ordinator**

Christian Medical College, Vellore

\*\*\*\*\* End of Report \*\*\*\*\*