

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN. 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE

| | | | |
|--------------------------|--|-----------------|-----------------|
| Customer Name & Address: | M/S. Priya Clinical Laboratory, 7/8A, North Car Street, (Chandhu), Thiruchendur – 628 215. | Certificate No: | AMS/LAB/2888818 |
| | | Issuing Date: | 15/12/2020 |
| | | Page No: | 1 of 1 |

Details of Device under Calibration (DUC):

| | | | |
|----------------------|---------------------------------|----------------------------|-----------------|
| DUC: | Centrifuge | Id. No: | PCL/LAB/CEN/001 |
| Make: | Remi | Department: | Laboratory |
| Model: | C – 852/4 | Calibrated On: | 14/12/2020 |
| Sl. No: | GNLC 4998 | Calibration Due Suggested: | 13/12/2021 |
| Type: | Manual 4BKTS | DUC Condition on Receipt: | OK |
| Range & Least Count: | 0 – 3500 RPM 0.5Amps 5 Steps | | |

Traceability Details:

| Sl.No. | Test Equipment | Make | Certificate Number | Valid Up to |
|--------|--------------------|-------------------|-------------------------|-------------|
| 1 | Digital Tachometer | Lutron / DT-2234C | FSMS/CL/05/19-07/203-01 | 29/07/2021 |
| 2 | Thermo Hygrometer | HTC | QC/AMS-DTI-004 | 22/02/2021 |

Room Temperature: 25°C±2

Room Humidity: 30-60%

(At the time of Calibration)

Test Results:

| SL.No. | DUC Reading (Position) | Standard Reading (RPM) | Correction (RPM) |
|--------|------------------------|------------------------|------------------|
| 1 | Position 1 | 1590 | 0 |
| 2 | Position 2 | 1882 | 0 |
| 3 | Position 3 | 2534 | 0 |
| 4 | Position 4 | 2922 | 0 |
| 5 | Position 5 | 3223 | 0 |

Allowable Tolerance: User Defined

Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test
(i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)

B. D. S.



Approved by:

(Technical Manager)

B. D. S.