

Ace Medical Solutions (India) Pvt. Ltd.

An ISO 9001-2015 Certified Company

No : 3/751, 1st Cross Street, Surveyor Colony, K.Pudur, Madurai - 625 007.

Phone : 0452 - 2561691 Mobile : +91 94434 74610



GSTIN: 33AAHCA8053H1ZR • DL No. MDU/4205/20B,MDU/3971/21B • CIN : U33110TN2009PTC072140

CALIBRATION CERTIFICATE

Customer Name & Address:	M/S. Priya Clinical Laboratory, 7/8A, North Car Street, (Chandhu), Thiruchendur - 628 215.	Certificate No:	AMS/LAB/2888824
		Issuing Date:	15/12/2020
		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Refrigerator - Reagent	Id. No:	PCL/LAB/REFR/001
Make:	LG	Department:	Laboratory
Model:	GL 205 KL4	Calibrated On:	14/12/2020
Sl. No:	2012 - 190	Calibration Due Suggested:	13/12/2021
Type:	Reagent Direct Cool	DUC Condition on Receipt:	OK
Range & Least Count:	0 - 8°C / LC 1°C		

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	RTD Thermometer	Center / RTD 375	QC/AMS/DTI/002	22/02/2021
2	Thermo Hygrometer	HTC	QC/AMS-DTI-004	22/02/2021

Room Temperature: 25°C±2 Room Humidity: 30-60%

(At the time of Calibration)

Test Results:

SL.No.	DUC Reading (°C)	Standard Reading (°C)	Correction (°C)
1	5.0	5.75	+0.75
2	-1.0	0.28	+1.28

Allowable Tolerance: User Defined

Note:

1. Calibration Points were selected as per customer request.
2. Each reading were taken by Repeatability Test (i.e.) each reading in a row is a mean of 10 readings.
3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)

B. B. S.



Approved by:

(Technical Manager)

B. B. S.