



Prima Calibration Services

F-199, MANGAL BAZAR LAXMI NAGAR DELHI-110092

Ph.: 011-42657237 Mob.: 9810804449, 8368671530

E- mail: primac21@gmail.com



Format No.
F01(7.8/01)

CALIBRATION CERTIFICATE

ULR No. : CC23392100002571F	Calibration Done On : 02.04.2021	S.R.F.No. & Date : 0002/02 & 02.04.2021
Issue Date : 03.04.2021	Calibration Valid Upto : 01.04.2022	Page : 1 of 1
UUC Receipt Date : 02.04.2021		

Calibrated for M/S.

Dr. Sardana Lab
Opposite Apeejay College,
Jalandhar City Punjab.

Identification on UUC (Unit Under Calibration)

Instrument Name	Micro Pipette	Size/Range	10 - 100 µl
Least Count	1 µl	Condition of Item	Good
Manufactured By	Dragon Lab	Calibration Performed at	Lab
Manufacturer M/S.No.	YE19BAM0061684	Location	--
Party ID Mark	--	Ref. Standard	ISO-8655-6

Calibration Procedure / Method : As per Calibration Procedure Document No. PCS/CP/01, Section No.CP-51

Environmental conditions (s) : Temperature: 23 ± 3 °C. Relative Humidity : 50 ± 10% RH

Standard & Major Equipment(s) Used :

Standard Equipment	Range/ Serial No.	Calibrating Agency	Traceability Cert. No.	Certificate Date	Valid Up to
Standard Weight	1 mg to 200 g	Weightronics	WMCL/E/2020-06/3369	03.06.2020	02.06.2023

Calibration Results at 27 °C :

Nominal Value UUC (µl)	Observed Value (µl)	Error (µl)	Related Uncertainty on C.F.(k=2) at approx 95% Confidence Level ±
10	10.42	-0.42	0.13 µl
50	50.51	-0.51	0.13 µl
100	100.94	-0.94	0.13 µl

***** END OF REPORT *****

NOTE:

- The result are valid only for this Calibrated item.
- The calibration report shall not be reproduced without the written approval of calibration laboratory.
- Result reported are valid at the time of & under the stated conditions of measurement
- Laboratory standards are traceable to National/International Standards.
- Calibration Certificates issued for Weight & Measure parameters i.e. Mass, Balance, Volumetric equipment, Measuring Scales / Tapes etc. are for scientific purpose only and should not be used for Trade / Commercial use.

CALIBRATED BY

shw

(CALIBRATION ENGINEER)



APPROVED BY

ANWAR A. SIDDIQUI

ANWAR A. SIDDIQUI
(Q.M/T.M.)